

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121025\
 Data File : VX048818.D
 Acq On : 10 Dec 2025 17:35
 Operator : JC/MD
 Sample : Q3776-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-20251203

Quant Time: Dec 11 00:27:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	137913	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	291436	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	314013	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	168146	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	103725	56.036	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 112.080%	
35) Dibromofluoromethane	5.379	113	90931	45.315	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 90.620%	
50) Toluene-d8	8.635	98	318234	45.243	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 90.480%#	
62) 4-Bromofluorobenzene	11.061	95	141105	58.175	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 116.360%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121025\
Data File : VX048818.D
Acq On : 10 Dec 2025 17:35
Operator : JC/MD
Sample : Q3776-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-20251203

Quant Time: Dec 11 00:27:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

