

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121118\
 Data File : VX006426.D
 Acq On : 11 Dec 2018 17:44
 Operator : JC/MD
 Sample : J6324-02DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-120D1-20181205DL

Quant Time: Dec 12 05:52:44 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	745862	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1142794	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	1013742	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	487501	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	371774	48.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.66%	
35) Dibromofluoromethane	5.50	113	349363	47.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.44%	
50) Toluene-d8	8.71	98	1324007	49.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.38%	
62) 4-Bromofluorobenzene	11.14	95	471323	46.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.94%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.38	101	11626	1.649	ug/l	97
12) 1,1-Dichloroethene	2.37	96	6345	0.940	ug/l	88
27) cis-1,2-Dichloroethene	4.59	96	2951	0.368	ug/l	89
44) Trichloroethene	7.21	130	709311	81.233	ug/l	97
64) Tetrachloroethene	9.34	164	2763	0.362	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121118\
Data File : VX006426.D
Acq On : 11 Dec 2018 17:44
Operator : JC/MD
Sample : J6324-02DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-120D1-20181205DL

Quant Time: Dec 12 05:52:44 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

