

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121118\
 Data File : VX006428.D
 Acq On : 11 Dec 2018 18:38
 Operator : JC/MD
 Sample : J6324-04DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-103D2-20181205DL

Quant Time: Dec 12 05:54:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	747056	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1144309	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	1031480	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	482302	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	377317	49.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.96%	
35) Dibromofluoromethane	5.51	113	351645	47.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.92%	
50) Toluene-d8	8.71	98	1333056	49.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.94%	
62) 4-Bromofluorobenzene	11.13	95	487330	47.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.98%	
Target Compounds						
44) Trichloroethene	7.21	130	769394	87.997	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121118\
Data File : VX006428.D
Acq On : 11 Dec 2018 18:38
Operator : JC/MD
Sample : J6324-04DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-103D2-20181205DL

Quant Time: Dec 12 05:54:47 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

