

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121119\
 Data File : VX013936.D
 Acq On : 11 Dec 2019 12:01
 Operator : JC/SP
 Sample : K6185-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP02-20191204

Quant Time: Dec 12 04:15:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	565692	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	872155	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	768941	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	343097	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	318751	48.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.08%	
35) Dibromofluoromethane	5.49	113	263035	48.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.56%	
50) Toluene-d8	8.71	98	1039015	49.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.92%	
62) 4-Bromofluorobenzene	11.14	95	339793	44.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.72%	
Target Compounds						
16) Acetone	2.43	43	11537	3.559	ug/l	93
44) Trichloroethene	7.21	130	160087	24.191	ug/l	96
64) Tetrachloroethene	9.33	164	5314	0.927	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121119\
Data File : VX013936.D
Acq On : 11 Dec 2019 12:01
Operator : JC/SP
Sample : K6185-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP02-20191204

Quant Time: Dec 12 04:15:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

