

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121119\
 Data File : VX013947.D
 Acq On : 11 Dec 2019 16:22
 Operator : JC/SP
 Sample : K6237-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY038PS037-191209

Quant Time: Dec 12 05:31:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	553813	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	861996	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	767510	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	328069	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	314414	48.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.82%	
35) Dibromofluoromethane	5.49	113	258635	48.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.06%	
50) Toluene-d8	8.71	98	1018536	49.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.12%	
62) 4-Bromofluorobenzene	11.14	95	338693	45.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.48%	
Target Compounds						
16) Acetone	2.43	43	31553	9.942	ug/l	97
25) 2-Butanone	4.68	43	14845	3.147	ug/l	95
27) cis-1,2-Dichloroethene	4.59	96	31966	4.848	ug/l	89
44) Trichloroethene	7.21	130	5103	0.780	ug/l	100
65) Chlorobenzene	10.14	112	6594	0.414	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121119\
Data File : VX013947.D
Acq On : 11 Dec 2019 16:22
Operator : JC/SP
Sample : K6237-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY038PS037-191209

Quant Time: Dec 12 05:31:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

