

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121119\
 Data File : VX013953.D
 Acq On : 11 Dec 2019 19:05
 Operator : JC/SP
 Sample : K6249-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-PZ188I-20191208

Quant Time: Dec 12 07:26:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	555783	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	861963	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	768061	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	340804	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	318843	49.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.84%	
35) Dibromofluoromethane	5.49	113	257273	48.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.56%	
50) Toluene-d8	8.71	98	1014092	48.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.68%	
62) 4-Bromofluorobenzene	11.13	95	337552	45.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.18%	
Target Compounds						
12) 1,1-Dichloroethene	2.36	96	8182	1.547	ug/l	96
24) 1,1-Dichloroethane	3.69	63	153314	15.174	ug/l	98
85) sec-Butylbenzene	11.94	105	10269	0.451	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121119\
Data File : VX013953.D
Acq On : 11 Dec 2019 19:05
Operator : JC/SP
Sample : K6249-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-PZ188I-20191208

Quant Time: Dec 12 07:26:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

