

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121119\
 Data File : VX013954.D
 Acq On : 11 Dec 2019 19:28
 Operator : JC/SP
 Sample : K6249-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-PZ189I-20191208

Quant Time: Dec 12 07:27:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	529392	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	815179	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	721587	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	306257	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	298730	48.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.22%	
35) Dibromofluoromethane	5.49	113	243396	48.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.60%	
50) Toluene-d8	8.71	98	964667	49.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.26%	
62) 4-Bromofluorobenzene	11.14	95	312856	44.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.38%	
Target Compounds						
24) 1,1-Dichloroethane	3.69	63	58180	6.045	ug/l	98
73) Isopropylbenzene	11.01	105	24956	1.162	ug/l	96
83) tert-Butylbenzene	11.76	119	9666	0.540	ug/l	85
85) sec-Butylbenzene	11.94	105	48467	2.371	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121119\
Data File : VX013954.D
Acq On : 11 Dec 2019 19:28
Operator : JC/SP
Sample : K6249-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-PZ189I-20191208

Quant Time: Dec 12 07:27:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

