

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121119\
 Data File : VX013963.D
 Acq On : 12 Dec 2019 00:07
 Operator : JC/SP
 Sample : K6252-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB02-20191209

Quant Time: Dec 12 06:39:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	543486	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	829365	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	744874	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	335653	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	302047	47.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.76%	
35) Dibromofluoromethane	5.49	113	241327	47.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.14%	
50) Toluene-d8	8.71	98	981081	49.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.22%	
62) 4-Bromofluorobenzene	11.14	95	338393	46.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.96%	
Target Compounds						
3) Chloromethane	1.31	50	7868	1.159	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121119\
Data File : VX013963.D
Acq On : 12 Dec 2019 00:07
Operator : JC/SP
Sample : K6252-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB02-20191209

Quant Time: Dec 12 06:39:21 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

