

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121120\
 Data File : VX019895.D
 Acq On : 11 Dec 2020 23:12
 Operator : JC/MD
 Sample : L4996-19DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D3-20201208DL

Quant Time: Dec 12 00:05:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	317984	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	533430	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.097	117	487534	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	210053	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	204239	48.87	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.74%
35) Dibromofluoromethane	5.464	113	166983	48.44	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.88%
50) Toluene-d8	8.695	98	643304	48.88	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.76%
62) 4-Bromofluorobenzene	11.122	95	227034	45.32	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.64%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.367	101	158718	55.08	ug/l	99
12) 1,1-Dichloroethene	2.361	96	1221	0.42	ug/l	79
44) Trichloroethene	7.189	130	55134	14.27	ug/l	99
49) 1,4-Dioxane	7.720	88	282	2.99	ug/l #	36
64) Tetrachloroethene	9.317	164	17216	5.27	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121120\
Data File : VX019895.D
Acq On : 11 Dec 2020 23:12
Operator : JC/MD
Sample : L4996-19DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20201208DL

Quant Time: Dec 12 00:05:34 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

