

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121120\
 Data File : VX019909.D
 Acq On : 12 Dec 2020 05:47
 Operator : JC/MD
 Sample : L5025-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TU032-32-EB-120920

Quant Time: Dec 12 06:53:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	319180	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	534433	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.098	117	494749	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	216099	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	204591	48.77	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.54%
35) Dibromofluoromethane	5.464	113	164270	47.56	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.12%
50) Toluene-d8	8.695	98	644886	48.91	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.82%
62) 4-Bromofluorobenzene	11.122	95	227957	45.42	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.84%
Target Compounds						
16) Acetone	2.416	43	2266	1.44	ug/l	97
17) Carbon Disulfide	2.550	76	2088	0.25	ug/l #	95
29) Tetrahydrofuran	5.086	42	28384	18.18	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121120\
Data File : VX019909.D
Acq On : 12 Dec 2020 05:47
Operator : JC/MD
Sample : L5025-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TU032-32-EB-120920

Quant Time: Dec 12 06:53:31 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

