

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX121120\
 Data File : VX019879.D
 Acq On : 11 Dec 2020 16:40
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 12 00:02:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	77	0.00
2 T	Dichlorodifluoromethane	50.000	48.793	2.4	70	0.00
3 P	Chloromethane	50.000	45.674	8.7	72	0.00
4 C	Vinyl Chloride	50.000	49.215	1.6#	73	0.00
5 T	Bromomethane	50.000	42.698	14.6	64	0.00
6 T	Chloroethane	50.000	53.449	-6.9	78	0.00
7 T	Trichlorofluoromethane	50.000	49.520	1.0	75	0.00
8 T	Diethyl Ether	50.000	53.648	-7.3	83	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	54.183	-8.4	82	0.00
10 T	Methyl Iodide	50.000	45.585	8.8	69	0.00
11 T	Tert butyl alcohol	250.000	258.220	-3.3	81	0.00
12 CM	1,1-Dichloroethene	50.000	51.455	-2.9#	77	0.00
13 T	Acrolein	250.000	218.295	12.7	70	0.00
14 T	Allyl chloride	50.000	53.159	-6.3	81	0.00
15 T	Acrylonitrile	250.000	283.016	-13.2	87	0.00
16 T	Acetone	250.000	257.947	-3.2	78	0.00
17 T	Carbon Disulfide	50.000	43.103	13.8	66	0.00
18 T	Methyl Acetate	50.000	59.186	-18.4	89	0.00
19 T	Methyl tert-butyl Ether	50.000	56.225	-12.5	85	0.00
20 T	Methylene Chloride	50.000	54.411	-8.8	81	0.00
21 T	trans-1,2-Dichloroethene	50.000	51.585	-3.2	78	0.00
22 T	Diisopropyl ether	50.000	57.068	-14.1	86	0.00
23 T	Vinyl Acetate	250.000	286.419	-14.6	85	0.00
24 P	1,1-Dichloroethane	50.000	53.939	-7.9	82	0.00
25 T	2-Butanone	250.000	279.321	-11.7	85	0.00
26 T	2,2-Dichloropropane	50.000	52.424	-4.8	79	0.00
27 T	cis-1,2-Dichloroethene	50.000	52.257	-4.5	81	0.00
28 T	Bromochloromethane	50.000	51.042	-2.1	78	0.00
29 T	Tetrahydrofuran	250.000	279.089	-11.6	86	0.00
30 C	Chloroform	50.000	54.767	-9.5#	83	0.00
31 T	Cyclohexane	50.000	50.103	-0.2	74	0.00
32 T	1,1,1-Trichloroethane	50.000	51.690	-3.4	77	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.062	1.9	79	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	78	0.00
35 S	Dibromofluoromethane	50.000	48.314	3.4	77	0.00
36 T	1,1-Dichloropropene	50.000	49.473	1.1	77	0.00
37 T	Ethyl Acetate	50.000	54.248	-8.5	86	0.00
38 T	Carbon Tetrachloride	50.000	49.391	1.2	75	0.00
39 T	Methylcyclohexane	50.000	50.370	-0.7	76	0.00
40 TM	Benzene	50.000	52.119	-4.2	81	0.00
41 T	Methacrylonitrile	50.000	54.746	-9.5	85	0.00
42 TM	1,2-Dichloroethane	50.000	54.128	-8.3	83	0.00
43 T	Isopropyl Acetate	50.000	54.346	-8.7	85	0.00
44 TM	Trichloroethene	50.000	50.352	-0.7	78	0.00
45 C	1,2-Dichloropropane	50.000	54.231	-8.5#	84	0.00
46 T	Dibromomethane	50.000	54.404	-8.8	84	0.00
47 T	Bromodichloromethane	50.000	54.479	-9.0	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX121120\
 Data File : VX019879.D
 Acq On : 11 Dec 2020 16:40
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 12 00:02:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48 T	Methyl methacrylate	50.000	54.910	-9.8	85	0.00
49 T	1,4-Dioxane	1000.000	1100.873	-10.1	86	0.00
50 S	Toluene-d8	50.000	48.256	3.5	76	0.00
51 T	4-Methyl-2-Pentanone	250.000	281.711	-12.7	87	0.00
52 CM	Toluene	50.000	52.203	-4.4#	80	0.00
53 T	t-1,3-Dichloropropene	50.000	53.702	-7.4	80	0.00
54 T	cis-1,3-Dichloropropene	50.000	54.904	-9.8	82	0.00
55 T	1,1,2-Trichloroethane	50.000	55.259	-10.5	85	0.00
56 T	Ethyl methacrylate	50.000	56.364	-12.7	85	0.00
57 T	1,3-Dichloropropane	50.000	54.421	-8.8	85	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	287.923	-15.2	89	0.00
59 T	2-Hexanone	250.000	283.612	-13.4	87	0.00
60 T	Dibromochloromethane	50.000	56.115	-12.2	84	0.00
61 T	1,2-Dibromoethane	50.000	54.151	-8.3	84	0.00
62 S	4-Bromofluorobenzene	50.000	49.696	0.6	78	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	79	0.00
64 T	Tetrachloroethene	50.000	54.668	-9.3	85	0.00
65 PM	Chlorobenzene	50.000	51.611	-3.2	80	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	55.049	-10.1	84	0.00
67 C	Ethyl Benzene	50.000	52.457	-4.9#	80	0.00
68 T	m/p-Xylenes	100.000	105.667	-5.7	80	0.00
69 T	o-Xylene	50.000	53.140	-6.3	81	0.00
70 T	Styrene	50.000	54.540	-9.1	81	0.00
71 P	Bromoform	50.000	54.209	-8.4	79	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	77	0.00
73 T	Isopropylbenzene	50.000	52.442	-4.9	80	0.00
74 T	N-amyl acetate	50.000	54.374	-8.7	82	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	54.196	-8.4	84	0.00
76 T	1,2,3-Trichloropropane	50.000	54.334	-8.7	84	0.00
77 T	Bromobenzene	50.000	51.683	-3.4	81	0.00
78 T	n-propylbenzene	50.000	52.558	-5.1	79	0.00
79 T	2-Chlorotoluene	50.000	52.616	-5.2	81	0.00
80 T	1,3,5-Trimethylbenzene	50.000	53.820	-7.6	80	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	55.017	-10.0	81	0.00
82 T	4-Chlorotoluene	50.000	52.787	-5.6	81	0.00
83 T	tert-Butylbenzene	50.000	52.166	-4.3	78	0.00
84 T	1,2,4-Trimethylbenzene	50.000	54.045	-8.1	80	0.00
85 T	sec-Butylbenzene	50.000	53.163	-6.3	79	0.00
86 T	p-Isopropyltoluene	50.000	53.570	-7.1	79	0.00
87 T	1,3-Dichlorobenzene	50.000	51.845	-3.7	79	0.00
88 T	1,4-Dichlorobenzene	50.000	51.990	-4.0	80	0.00
89 T	n-Butylbenzene	50.000	53.670	-7.3	79	0.00
90 T	Hexachloroethane	50.000	52.690	-5.4	79	0.00
91 T	1,2-Dichlorobenzene	50.000	54.243	-8.5	82	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	53.762	-7.5	83	0.00
93 T	1,2,4-Trichlorobenzene	50.000	54.351	-8.7	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121120\
Data File : VX019879.D
Acq On : 11 Dec 2020 16:40
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 12 00:02:10 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : Sw846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
94 T	Hexachlorobutadiene	50.000	53.123	-6.2	84	0.00
95 T	Naphthalene	50.000	57.130	-14.3	84	0.00
96 T	1,2,3-Trichlorobenzene	50.000	54.994	-10.0	82	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6