

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121120\
 Data File : VX019902.D
 Acq On : 12 Dec 2020 01:55
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 12 02:51:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	76	0.00
2 T	Dichlorodifluoromethane	0.458	0.436	4.8	68	0.00
3 P	Chloromethane	0.538	0.504	6.3	73	0.00
4 C	Vinyl Chloride	0.568	0.569	-0.2#	74	0.00
5 T	Bromomethane	0.188	0.172	8.5	68	0.00
6 T	Chloroethane	0.341	0.378	-10.9	80	0.00
7 T	Trichlorofluoromethane	0.791	0.786	0.6	74	0.00
8 T	Diethyl Ether	0.335	0.352	-5.1	80	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.453	0.485	-7.1	79	0.00
10 T	Methyl Iodide	0.668	0.615	7.9	69	0.00
11 T	Tert butyl alcohol	0.145	0.135	6.9	77	0.00
12 CM	1,1-Dichloroethene	0.458	0.478	-4.4#	77	0.00
13 T	Acrolein	0.073	0.055	24.7	60	0.00
14 T	Allyl chloride	0.774	0.805	-4.0	79	0.00
15 T	Acrylonitrile	0.297	0.330	-11.1	85	0.00
16 T	Acetone	0.247	0.257	-4.0	78	0.00
17 T	Carbon Disulfide	1.314	1.147	12.7	66	0.00
18 T	Methyl Acetate	0.604	0.726	-20.2	90	0.00
19 T	Methyl tert-butyl Ether	1.668	1.859	-11.5	83	0.00
20 T	Methylene Chloride	0.639	0.614	3.9	80	0.00
21 T	trans-1,2-Dichloroethene	0.534	0.553	-3.6	77	0.00
22 T	Diisopropyl ether	1.564	1.780	-13.8	84	0.00
23 T	Vinyl Acetate	1.333	1.502	-12.7	82	0.00
24 P	1,1-Dichloroethane	0.943	1.030	-9.2	82	0.00
25 T	2-Butanone	0.380	0.417	-9.7	83	0.00
26 T	2,2-Dichloropropane	0.812	0.737	9.2	68	0.00
27 T	cis-1,2-Dichloroethene	0.625	0.664	-6.2	81	0.00
28 T	Bromochloromethane	0.450	0.465	-3.3	78	0.00
29 T	Tetrahydrofuran	0.245	0.269	-9.8	84	0.00
30 C	Chloroform	0.970	1.065	-9.8#	82	0.00
31 T	Cyclohexane	0.817	0.828	-1.3	74	0.00
32 T	1,1,1-Trichloroethane	0.835	0.885	-6.0	78	0.00
33 S	1,2-Dichloroethane-d4	0.657	0.655	0.3	80	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	78	0.00
35 S	Dibromofluoromethane	0.323	0.315	2.5	78	0.00
36 T	1,1-Dichloropropene	0.452	0.443	2.0	76	0.00
37 T	Ethyl Acetate	0.456	0.481	-5.5	84	0.00
38 T	Carbon Tetrachloride	0.442	0.437	1.1	75	0.00
39 T	Methylcyclohexane	0.531	0.510	4.0	73	0.00
40 TM	Benzene	1.380	1.422	-3.0	80	0.00
41 T	Methacrylonitrile	0.250	0.264	-5.6	82	0.00
42 TM	1,2-Dichloroethane	0.472	0.495	-4.9	81	0.00
43 T	Isopropyl Acetate	0.742	0.777	-4.7	82	0.00
44 TM	Trichloroethene	0.362	0.355	1.9	76	0.00
45 C	1,2-Dichloropropane	0.345	0.371	-7.5#	84	0.00
46 T	Dibromomethane	0.233	0.245	-5.2	81	0.00
47 T	Bromodichloromethane	0.462	0.487	-5.4	80	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121120\
 Data File : VX019902.D
 Acq On : 12 Dec 2020 01:55
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 12 02:51:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48 T	Methyl methacrylate	0.365	0.389	-6.6	83	0.00
49 T	1,4-Dioxane	0.009	0.009	0.0	77	0.00
50 S	Toluene-d8	1.234	1.202	2.6	77	0.00
51 T	4-Methyl-2-Pentanone	0.453	0.494	-9.1	85	0.00
52 CM	Toluene	0.863	0.892	-3.4#	79	0.00
53 T	t-1,3-Dichloropropene	0.513	0.526	-2.5	76	0.00
54 T	cis-1,3-Dichloropropene	0.557	0.575	-3.2	77	0.00
55 T	1,1,2-Trichloroethane	0.348	0.375	-7.8	83	0.00
56 T	Ethyl methacrylate	0.526	0.574	-9.1	82	0.00
57 T	1,3-Dichloropropane	0.594	0.629	-5.9	83	0.00
58 T	2-Chloroethyl Vinyl ether	0.279	0.310	-11.1	86	0.00
59 T	2-Hexanone	0.351	0.374	-6.6	82	0.00
60 T	Dibromochloromethane	0.353	0.380	-7.6	80	0.00
61 T	1,2-Dibromoethane	0.369	0.385	-4.3	81	0.00
62 S	4-Bromofluorobenzene	0.470	0.460	2.1	77	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	80	0.00
64 T	Tetrachloroethene	0.335	0.322	3.9	76	0.00
65 PM	Chlorobenzene	0.996	0.983	1.3	78	0.00
66 T	1,1,1,2-Tetrachloroethane	0.359	0.369	-2.8	80	0.00
67 C	Ethyl Benzene	1.768	1.781	-0.7#	78	0.00
68 T	m/p-Xylenes	0.666	0.676	-1.5	79	0.00
69 T	o-Xylene	0.641	0.653	-1.9	79	0.00
70 T	Styrene	1.082	1.127	-4.2	79	0.00
71 P	Bromoform	0.282	0.291	-3.2	77	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	76	0.00
73 T	Isopropylbenzene	3.416	3.531	-3.4	79	0.00
74 T	N-amyl acetate	1.369	1.462	-6.8	80	0.00
75 P	1,1,2,2-Tetrachloroethane	1.204	1.284	-6.6	82	0.00
76 T	1,2,3-Trichloropropane	1.121	1.171	-4.5	80	0.00
77 T	Bromobenzene	0.874	0.891	-1.9	80	0.00
78 T	n-propylbenzene	3.902	4.045	-3.7	78	0.00
79 T	2-Chlorotoluene	2.403	2.471	-2.8	79	0.00
80 T	1,3,5-Trimethylbenzene	2.883	3.017	-4.6	77	0.00
81 T	trans-1,4-Dichloro-2-butene	0.388	0.384	1.0	73	0.00
82 T	4-Chlorotoluene	2.814	2.890	-2.7	79	0.00
83 T	tert-Butylbenzene	2.828	2.950	-4.3	78	0.00
84 T	1,2,4-Trimethylbenzene	2.904	3.081	-6.1	78	0.00
85 T	sec-Butylbenzene	3.228	3.381	-4.7	78	0.00
86 T	p-Isopropyltoluene	2.959	3.121	-5.5	77	0.00
87 T	1,3-Dichlorobenzene	1.604	1.607	-0.2	76	0.00
88 T	1,4-Dichlorobenzene	1.634	1.604	1.8	76	0.00
89 T	n-Butylbenzene	2.621	2.682	-2.3	75	0.00
90 T	Hexachloroethane	0.511	0.538	-5.3	79	0.00
91 T	1,2-Dichlorobenzene	1.564	1.627	-4.0	79	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.259	0.266	-2.7	79	0.00
93 T	1,2,4-Trichlorobenzene	1.018	1.069	-5.0	78	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121120\
Data File : VX019902.D
Acq On : 12 Dec 2020 01:55
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 12 02:51:02 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : Sw846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
94 T	Hexachlorobutadiene	0.455	0.466	-2.4	80	0.00
95 T	Naphthalene	3.301	3.592	-8.8	80	0.00
96 T	1,2,3-Trichlorobenzene	0.995	1.054	-5.9	79	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6