

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121120\
 Data File : VX019904.D
 Acq On : 12 Dec 2020 03:04
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 12 05:38:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	76	0.00
2 T	Dichlorodifluoromethane	50.000	49.266	1.5	70	0.00
3 P	Chloromethane	50.000	48.121	3.8	75	0.00
4 C	Vinyl Chloride	50.000	51.298	-2.6#	76	0.00
5 T	Bromomethane	50.000	46.452	7.1	69	0.00
6 T	Chloroethane	50.000	56.002	-12.0	81	0.00
7 T	Trichlorofluoromethane	50.000	51.861	-3.7	78	0.00
8 T	Diethyl Ether	50.000	52.590	-5.2	81	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	55.260	-10.5	82	0.00
10 T	Methyl Iodide	50.000	47.836	4.3	71	0.00
11 T	Tert butyl alcohol	250.000	267.732	-7.1	83	0.00
12 CM	1,1-Dichloroethene	50.000	54.119	-8.2#	81	0.00
13 T	Acrolein	250.000	186.227	25.5#	59	0.00
14 T	Allyl chloride	50.000	52.619	-5.2	80	0.00
15 T	Acrylonitrile	250.000	284.010	-13.6	87	0.00
16 T	Acetone	250.000	269.628	-7.9	81	0.00
17 T	Carbon Disulfide	50.000	44.821	10.4	68	0.00
18 T	Methyl Acetate	50.000	60.031	-20.1	90	0.00
19 T	Methyl tert-butyl Ether	50.000	55.293	-10.6	82	0.00
20 T	Methylene Chloride	50.000	55.019	-10.0	81	0.00
21 T	trans-1,2-Dichloroethene	50.000	52.047	-4.1	78	0.00
22 T	Diisopropyl ether	50.000	56.887	-13.8	85	0.00
23 T	Vinyl Acetate	250.000	281.736	-12.7	83	0.00
24 P	1,1-Dichloroethane	50.000	54.777	-9.6	83	0.00
25 T	2-Butanone	250.000	282.234	-12.9	85	0.00
26 T	2,2-Dichloropropane	50.000	45.442	9.1	68	0.00
27 T	cis-1,2-Dichloroethene	50.000	52.193	-4.4	80	0.00
28 T	Bromochloromethane	50.000	51.330	-2.7	78	0.00
29 T	Tetrahydrofuran	250.000	282.610	-13.0	86	0.00
30 C	Chloroform	50.000	54.722	-9.4#	82	0.00
31 T	Cyclohexane	50.000	52.347	-4.7	77	0.00
32 T	1,1,1-Trichloroethane	50.000	53.778	-7.6	80	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.365	1.3	79	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	77	0.00
35 S	Dibromofluoromethane	50.000	49.188	1.6	79	0.00
36 T	1,1-Dichloropropene	50.000	50.405	-0.8	78	0.00
37 T	Ethyl Acetate	50.000	53.862	-7.7	85	0.00
38 T	Carbon Tetrachloride	50.000	51.571	-3.1	78	0.00
39 T	Methylcyclohexane	50.000	50.197	-0.4	75	0.00
40 TM	Benzene	50.000	52.138	-4.3	81	0.00
41 T	Methacrylonitrile	50.000	53.901	-7.8	84	0.00
42 TM	1,2-Dichloroethane	50.000	53.122	-6.2	81	0.00
43 T	Isopropyl Acetate	50.000	53.081	-6.2	83	0.00
44 TM	Trichloroethene	50.000	50.155	-0.3	77	0.00
45 C	1,2-Dichloropropane	50.000	53.577	-7.2#	83	0.00
46 T	Dibromomethane	50.000	52.840	-5.7	81	0.00
47 T	Bromodichloromethane	50.000	53.710	-7.4	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX121120\
 Data File : VX019904.D
 Acq On : 12 Dec 2020 03:04
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 12 05:38:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48 T	Methyl methacrylate	50.000	53.900	-7.8	83	0.00
49 T	1,4-Dioxane	1000.000	1085.013	-8.5	84	0.00
50 S	Toluene-d8	50.000	48.622	2.8	76	0.00
51 T	4-Methyl-2-Pentanone	250.000	281.398	-12.6	87	0.00
52 CM	Toluene	50.000	52.219	-4.4#	80	0.00
53 T	t-1,3-Dichloropropene	50.000	51.473	-2.9	76	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.236	-4.5	78	0.00
55 T	1,1,2-Trichloroethane	50.000	54.624	-9.2	84	0.00
56 T	Ethyl methacrylate	50.000	54.926	-9.9	82	0.00
57 T	1,3-Dichloropropane	50.000	53.694	-7.4	83	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	278.465	-11.4	86	0.00
59 T	2-Hexanone	250.000	276.681	-10.7	85	0.00
60 T	Dibromochloromethane	50.000	53.729	-7.5	80	0.00
61 T	1,2-Dibromoethane	50.000	53.070	-6.1	82	0.00
62 S	4-Bromofluorobenzene	50.000	49.983	0.0	78	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	79	0.00
64 T	Tetrachloroethene	50.000	54.200	-8.4	84	0.00
65 PM	Chlorobenzene	50.000	51.137	-2.3	80	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	54.107	-8.2	83	0.00
67 C	Ethyl Benzene	50.000	51.690	-3.4#	79	0.00
68 T	m/p-Xylenes	100.000	103.577	-3.6	79	0.00
69 T	o-Xylene	50.000	53.034	-6.1	81	0.00
70 T	Styrene	50.000	53.383	-6.8	79	0.00
71 P	Bromoform	50.000	53.013	-6.0	78	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	75	0.00
73 T	Isopropylbenzene	50.000	53.389	-6.8	80	0.00
74 T	N-amyl acetate	50.000	54.245	-8.5	80	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	54.713	-9.4	83	0.00
76 T	1,2,3-Trichloropropane	50.000	54.358	-8.7	82	0.00
77 T	Bromobenzene	50.000	52.852	-5.7	81	0.00
78 T	n-propylbenzene	50.000	53.487	-7.0	79	0.00
79 T	2-Chlorotoluene	50.000	52.254	-4.5	79	0.00
80 T	1,3,5-Trimethylbenzene	50.000	53.503	-7.0	78	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	50.909	-1.8	74	0.00
82 T	4-Chlorotoluene	50.000	52.820	-5.6	80	0.00
83 T	tert-Butylbenzene	50.000	52.510	-5.0	77	0.00
84 T	1,2,4-Trimethylbenzene	50.000	54.120	-8.2	78	0.00
85 T	sec-Butylbenzene	50.000	54.621	-9.2	80	0.00
86 T	p-Isopropyltoluene	50.000	53.886	-7.8	78	0.00
87 T	1,3-Dichlorobenzene	50.000	50.039	-0.1	75	0.00
88 T	1,4-Dichlorobenzene	50.000	50.643	-1.3	77	0.00
89 T	n-Butylbenzene	50.000	53.707	-7.4	78	0.00
90 T	Hexachloroethane	50.000	53.900	-7.8	80	0.00
91 T	1,2-Dichlorobenzene	50.000	52.782	-5.6	79	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	54.834	-9.7	83	0.00
93 T	1,2,4-Trichlorobenzene	50.000	52.694	-5.4	77	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121120\
Data File : VX019904.D
Acq On : 12 Dec 2020 03:04
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 12 05:38:49 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : Sw846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
94 T	Hexachlorobutadiene	50.000	52.795	-5.6	82	0.00
95 T	Naphthalene	50.000	56.972	-13.9	83	0.00
96 T	1,2,3-Trichlorobenzene	50.000	55.721	-11.4	81	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6