

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121120\
 Data File : VX019924.D
 Acq On : 12 Dec 2020 11:35
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 64 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 14 02:49:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	76	0.00
2 T	Dichlorodifluoromethane	50.000	48.855	2.3	70	0.00
3 P	Chloromethane	50.000	48.857	2.3	77	0.00
4 C	Vinyl Chloride	50.000	51.726	-3.5#	76	0.00
5 T	Bromomethane	50.000	47.819	4.4	71	0.00
6 T	Chloroethane	50.000	55.826	-11.7	81	0.00
7 T	Trichlorofluoromethane	50.000	51.898	-3.8	78	0.00
8 T	Diethyl Ether	50.000	52.712	-5.4	81	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	54.038	-8.1	81	0.00
10 T	Methyl Iodide	50.000	48.213	3.6	72	0.00
11 T	Tert butyl alcohol	250.000	282.831	-13.1	87	0.00
12 CM	1,1-Dichloroethene	50.000	54.578	-9.2#	81	0.00
13 T	Acrolein	250.000	206.965	17.2	66	0.00
14 T	Allyl chloride	50.000	51.961	-3.9	79	0.00
15 T	Acrylonitrile	250.000	285.440	-14.2	87	0.00
16 T	Acetone	250.000	270.584	-8.2	82	0.00
17 T	Carbon Disulfide	50.000	45.887	8.2	70	0.00
18 T	Methyl Acetate	50.000	60.420	-20.8	90	0.00
19 T	Methyl tert-butyl Ether	50.000	55.629	-11.3	83	0.00
20 T	Methylene Chloride	50.000	54.674	-9.3	80	0.00
21 T	trans-1,2-Dichloroethene	50.000	52.286	-4.6	78	0.00
22 T	Diisopropyl ether	50.000	57.184	-14.4	85	0.00
23 T	Vinyl Acetate	250.000	281.953	-12.8	83	0.00
24 P	1,1-Dichloroethane	50.000	54.996	-10.0	83	0.00
25 T	2-Butanone	250.000	284.881	-14.0	86	0.00
26 T	2,2-Dichloropropane	50.000	38.385	23.2	57	0.00
27 T	cis-1,2-Dichloroethene	50.000	53.022	-6.0	81	0.00
28 T	Bromochloromethane	50.000	52.436	-4.9	79	0.00
29 T	Tetrahydrofuran	250.000	284.881	-14.0	87	0.00
30 C	Chloroform	50.000	55.528	-11.1#	83	0.00
31 T	Cyclohexane	50.000	52.175	-4.3	77	0.00
32 T	1,1,1-Trichloroethane	50.000	54.640	-9.3	81	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.465	1.1	79	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	78	0.00
35 S	Dibromofluoromethane	50.000	49.097	1.8	79	0.00
36 T	1,1-Dichloropropene	50.000	49.458	1.1	77	0.00
37 T	Ethyl Acetate	50.000	54.026	-8.1	87	0.00
38 T	Carbon Tetrachloride	50.000	51.993	-4.0	79	0.00
39 T	Methylcyclohexane	50.000	48.199	3.6	73	0.00
40 TM	Benzene	50.000	51.933	-3.9	81	0.00
41 T	Methacrylonitrile	50.000	55.012	-10.0	86	0.00
42 TM	1,2-Dichloroethane	50.000	52.999	-6.0	82	0.00
43 T	Isopropyl Acetate	50.000	52.983	-6.0	84	0.00
44 TM	Trichloroethene	50.000	50.222	-0.4	78	0.00
45 C	1,2-Dichloropropane	50.000	53.938	-7.9#	84	0.00
46 T	Dibromomethane	50.000	52.903	-5.8	82	0.00
47 T	Bromodichloromethane	50.000	53.843	-7.7	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX121120\
 Data File : VX019924.D
 Acq On : 12 Dec 2020 11:35
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 64 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 14 02:49:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48 T	Methyl methacrylate	50.000	53.807	-7.6	84	0.00
49 T	1,4-Dioxane	1000.000	1085.490	-8.5	85	0.00
50 S	Toluene-d8	50.000	48.181	3.6	76	0.00
51 T	4-Methyl-2-Pentanone	250.000	280.405	-12.2	87	0.00
52 CM	Toluene	50.000	52.007	-4.0#	80	0.00
53 T	t-1,3-Dichloropropene	50.000	50.620	-1.2	76	0.00
54 T	cis-1,3-Dichloropropene	50.000	50.601	-1.2	76	0.00
55 T	1,1,2-Trichloroethane	50.000	54.688	-9.4	85	0.00
56 T	Ethyl methacrylate	50.000	55.080	-10.2	83	0.00
57 T	1,3-Dichloropropane	50.000	53.127	-6.3	83	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	277.639	-11.1	86	0.00
59 T	2-Hexanone	250.000	278.808	-11.5	86	0.00
60 T	Dibromochloromethane	50.000	54.568	-9.1	82	0.00
61 T	1,2-Dibromoethane	50.000	52.856	-5.7	83	0.00
62 S	4-Bromofluorobenzene	50.000	48.881	2.2	78	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	79	0.00
64 T	Tetrachloroethene	50.000	51.563	-3.1	81	0.00
65 PM	Chlorobenzene	50.000	50.612	-1.2	79	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	53.067	-6.1	82	0.00
67 C	Ethyl Benzene	50.000	51.928	-3.9#	80	0.00
68 T	m/p-Xylenes	100.000	103.441	-3.4	79	0.00
69 T	o-Xylene	50.000	51.924	-3.8	80	0.00
70 T	Styrene	50.000	53.345	-6.7	80	0.00
71 P	Bromoform	50.000	53.613	-7.2	79	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	76	0.00
73 T	Isopropylbenzene	50.000	53.162	-6.3	81	0.00
74 T	N-amyl acetate	50.000	53.934	-7.9	81	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	54.519	-9.0	84	0.00
76 T	1,2,3-Trichloropropane	50.000	53.131	-6.3	81	0.00
77 T	Bromobenzene	50.000	52.181	-4.4	81	0.00
78 T	n-propylbenzene	50.000	52.231	-4.5	78	0.00
79 T	2-Chlorotoluene	50.000	52.298	-4.6	80	0.00
80 T	1,3,5-Trimethylbenzene	50.000	52.269	-4.5	77	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	47.858	4.3	70	0.00
82 T	4-Chlorotoluene	50.000	51.624	-3.2	79	0.00
83 T	tert-Butylbenzene	50.000	52.655	-5.3	79	0.00
84 T	1,2,4-Trimethylbenzene	50.000	53.066	-6.1	78	0.00
85 T	sec-Butylbenzene	50.000	53.018	-6.0	78	0.00
86 T	p-Isopropyltoluene	50.000	53.119	-6.2	78	0.00
87 T	1,3-Dichlorobenzene	50.000	50.259	-0.5	76	0.00
88 T	1,4-Dichlorobenzene	50.000	50.381	-0.8	78	0.00
89 T	n-Butylbenzene	50.000	51.111	-2.2	75	0.00
90 T	Hexachloroethane	50.000	53.149	-6.3	80	0.00
91 T	1,2-Dichlorobenzene	50.000	52.149	-4.3	79	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	54.225	-8.5	84	0.00
93 T	1,2,4-Trichlorobenzene	50.000	52.265	-4.5	78	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121120\
Data File : VX019924.D
Acq On : 12 Dec 2020 11:35
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 64 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 14 02:49:18 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : Sw846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
94 T	Hexachlorobutadiene	50.000	52.710	-5.4	83	0.00
95 T	Naphthalene	50.000	55.507	-11.0	82	0.00
96 T	1,2,3-Trichlorobenzene	50.000	53.206	-6.4	79	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6