



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121120\
 Data File : VX019864.D
 Acq On : 11 Dec 2020 08:42
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Integration File: RTEINT3.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\624X120820W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Wed Dec 09 06:09:21 2020



AutoFind: Averaged scan 1640 to 1658; Bkg corrected with scan 1639

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.8	2430	PASS
75	95	30	60	48.0	6934	PASS
95	95	100	100	100.0	14443	PASS
96	95	5	9	6.9	991	PASS
173	174	0.00	2	0.7	85	PASS
174	95	50	100	79.0	11405	PASS
175	174	5	9	7.1	811	PASS
176	174	95	101	95.6	10903	PASS
177	176	5	9	6.5	709	PASS