



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121120\
Data File : VX019864.D
Acq On : 11 Dec 2020 08:42
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Title : SW846 8260
Last Update : Fri Dec 04 15:32:53 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1638

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	16.0	10323	PASS
75	95	30	60	47.1	30357	PASS
95	95	100	100	100.0	64509	PASS
96	95	5	9	6.6	4257	PASS
173	174	0.00	2	0.8	424	PASS
174	95	50	100	82.1	52979	PASS
175	174	5	9	7.2	3796	PASS
176	174	95	101	95.5	50621	PASS
177	176	5	9	6.4	3262	PASS