

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121123\
 Data File : VX039130.D
 Acq On : 11 Dec 2023 10:01
 Operator : JC/MD
 Sample : VX1211MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK644

Quant Time: Dec 12 03:51:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML120523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 12 03:50:51 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	268992	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	246045	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	118743	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	87842	41.635	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.260%
7) Chloroethane-d5	1.672	69	66484	42.632	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	85.260%
11) 1,1-Dichloroethene-d2	2.306	65	41879	45.219	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.440%
21) 2-Butanone-d5	4.446	46	190002	99.487	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.490%
24) Chloroform-d	5.056	84	173587	46.477	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.960%
26) 1,2-Dichloroethane-d4	5.952	65	118207	48.870	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.740%
32) Benzene-d6	5.970	84	355582	46.827	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.660%
36) 1,2-Dichloropropane-d6	7.306	67	118876	48.646	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.300%
41) Toluene-d8	8.647	98	311301	45.238	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.480%
43) trans-1,3-Dichloroprop...	8.946	79	57219	46.012	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.020%
47) 2-Hexanone-d5	9.378	63	141951	97.256	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.260%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	168105	48.492	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.980%
66) 1,2-Dichlorobenzene-d4	12.317	152	107968	48.104	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121123\
Data File : VX039130.D
Acq On : 11 Dec 2023 10:01
Operator : JC/MD
Sample : VX1211MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK644

Quant Time: Dec 12 03:51:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM120523WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 12 03:50:51 2023
Response via : Initial Calibration

