

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121124\
 Data File : VX044223.D
 Acq On : 11 Dec 2024 12:37
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC100

Quant Time: Dec 12 04:15:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:11:59 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 12/12/2024
 Supervised By : Mahesh Dadoda 12/12/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	130101	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	235738	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	209164	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97670	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	224292	98.326	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 196.660%#			
35) Dibromofluoromethane	5.379	113	169582	103.866	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 207.740%#			
50) Toluene-d8	8.647	98	579477	105.201	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 210.400%#			
62) 4-Bromofluorobenzene	11.079	95	203694	105.830	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 211.660%#			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.167	85	193785	96.329	ug/l	97
3) Chloromethane	1.295	50	186167	96.774	ug/l	99
4) Vinyl Chloride	1.374	62	187739	93.675	ug/l	99
5) Bromomethane	1.593	94	137770	96.897	ug/l	97
6) Chloroethane	1.660	64	109912	88.074	ug/l	95
7) Trichlorofluoromethane	1.862	101	384622	98.030	ug/l	98
8) Diethyl Ether	2.136	74	124039	92.618	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.313	101	149151	93.944	ug/l	99
10) Methyl Iodide	2.441	142	203545	97.275	ug/l	100
11) Tert butyl alcohol	3.008	59	173139	539.406	ug/l	100
12) 1,1-Dichloroethene	2.307	96	144268	98.930	ug/l	100
13) Acrolein	2.240	56	226037	485.718	ug/l	99
14) Allyl chloride	2.654	41	244602	100.703	ug/l	99
15) Acrylonitrile	3.069	53	446747	501.661	ug/l	99
16) Acetone	2.392	43	452955	487.065	ug/l	97
17) Carbon Disulfide	2.496	76	320036	115.453	ug/l	99
18) Methyl Acetate	2.709	43	230998	96.338	ug/l	97
19) Methyl tert-butyl Ether	3.117	73	564545	99.205	ug/l	99
20) Methylene Chloride	2.782	84	168498	96.466	ug/l	97
21) trans-1,2-Dichloroethene	3.081	96	154797	99.880	ug/l	98
22) Diisopropyl ether	3.764	45	538742	98.498	ug/l	99
23) Vinyl Acetate	3.721	43	2395923	538.801	ug/l	100
24) 1,1-Dichloroethane	3.605	63	304626	99.915	ug/l	99
25) 2-Butanone	4.568	43	662768	501.306	ug/l	99
26) 2,2-Dichloropropane	4.471	77	258092	104.201	ug/l	99
27) cis-1,2-Dichloroethene	4.483	96	190127	96.593	ug/l	99
28) Bromochloromethane	4.891	49	144783	96.842	ug/l	99
29) Tetrahydrofuran	5.007	42	412666	486.290	ug/l	99
30) Chloroform	5.093	83	331047	96.642	ug/l	98
31) Cyclohexane	5.458	56	239879	97.433	ug/l	98
32) 1,1,1-Trichloroethane	5.379	97	291065	102.363	ug/l	99
36) 1,1-Dichloropropene	5.684	75	209732	99.376	ug/l	98
37) Ethyl Acetate	4.721	43	266307	102.079	ug/l	100
38) Carbon Tetrachloride	5.672	117	236869	104.249	ug/l	96
39) Methylcyclohexane	7.373	83	259368	102.800	ug/l	98
40) Benzene	6.032	78	638055	99.575	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121124\
 Data File : VX044223.D
 Acq On : 11 Dec 2024 12:37
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC100

Quant Time: Dec 12 04:15:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:11:59 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/12/2024
 Supervised By : Mahesh Dadoda 12/12/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	140375	101.332	ug/l	99
42) 1,2-Dichloroethane	6.086	62	261991	99.255	ug/l	99
43) Isopropyl Acetate	6.342	43	427893	104.939	ug/l	99
44) Trichloroethene	7.123	130	158990	99.497	ug/l	99
45) 1,2-Dichloropropane	7.428	63	160913	98.293	ug/l	97
46) Dibromomethane	7.580	93	130316	102.607	ug/l	99
47) Bromodichloromethane	7.818	83	244401	111.423	ug/l	100
48) Methyl methacrylate	7.696	41	203116	102.618	ug/l	99
49) 1,4-Dioxane	7.665	88	73434m	2170.871	ug/l	
51) 4-Methyl-2-Pentanone	8.574	43	1322678	504.496	ug/l	99
52) Toluene	8.714	92	396383	98.506	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	248092	99.321	ug/l	95
54) cis-1,3-Dichloropropene	8.366	75	263014	110.127	ug/l	98
55) 1,1,2-Trichloroethane	9.153	97	157366	98.441	ug/l	98
56) Ethyl methacrylate	9.116	69	271146	107.442	ug/l	99
57) 1,3-Dichloropropane	9.305	76	274642	98.704	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.238	63	671233	527.342	ug/l	100
59) 2-Hexanone	9.433	43	989466	521.401	ug/l	99
60) Dibromochloromethane	9.519	129	178135	99.480	ug/l	99
61) 1,2-Dibromoethane	9.610	107	171458	105.371	ug/l	97
64) Tetrachloroethene	9.269	164	130678	94.131	ug/l	97
65) Chlorobenzene	10.080	112	433179	96.650	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.159	131	153882	106.831	ug/l	99
67) Ethyl Benzene	10.195	91	765930	98.906	ug/l	97
68) m/p-Xylenes	10.299	106	571648	201.171	ug/l	99
69) o-Xylene	10.640	106	280603	97.597	ug/l	100
70) Styrene	10.653	104	480077	104.159	ug/l	97
71) Bromoform	10.799	173	112482	99.723	ug/l #	98
73) Isopropylbenzene	10.964	105	731955	97.894	ug/l	99
74) N-amyl acetate	10.842	43	371420	107.485	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.214	83	245395	94.491	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	206786m	93.302	ug/l	
77) Bromobenzene	11.195	156	175647	96.874	ug/l	98
78) n-propylbenzene	11.305	91	836676	101.242	ug/l	99
79) 2-Chlorotoluene	11.366	91	504407	94.377	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	604993	97.570	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	78453	97.165	ug/l	94
82) 4-Chlorotoluene	11.451	91	588197	99.384	ug/l	99
83) tert-Butylbenzene	11.713	119	606738	99.013	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	614999	100.994	ug/l	100
85) sec-Butylbenzene	11.890	105	743098	99.889	ug/l	100
86) p-Isopropyltoluene	12.006	119	624416	100.847	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	317377	100.210	ug/l	99
88) 1,4-Dichlorobenzene	12.043	146	316688	97.271	ug/l	99
89) n-Butylbenzene	12.329	91	564303	108.385	ug/l	99
90) Hexachloroethane	12.536	117	105946	117.218	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	323040	98.232	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	54588	108.279	ug/l	93
93) 1,2,4-Trichlorobenzene	13.585	180	197899	109.766	ug/l	98
94) Hexachlorobutadiene	13.725	225	72453	99.908	ug/l	97
95) Naphthalene	13.774	128	755181	107.308	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	206053	107.754	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121124\
Data File : VX044223.D
Acq On : 11 Dec 2024 12:37
Operator : JC/MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :John Carlone 12/12/2024
Supervised By :Mahesh Dadoda 12/12/2024

Quant Time: Dec 12 04:15:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:11:59 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121124\
Data File : VX044223.D
Acq On : 11 Dec 2024 12:37
Operator : JC/MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Quant Time: Dec 12 04:15:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:11:59 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 12/12/2024
Supervised By :Mahesh Dadoda 12/12/2024

