

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121124\
 Data File : VX044228.D
 Acq On : 11 Dec 2024 14:51
 Operator : JC/MD
 Sample : VX1211WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1211WBL01

Quant Time: Dec 12 04:52:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	110553	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	207554	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	181474	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74486	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	95794	49.420	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.840%
35) Dibromofluoromethane	5.385	113	68679	47.777	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.560%
50) Toluene-d8	8.647	98	250096	51.569	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.140%
62) 4-Bromofluorobenzene	11.079	95	83078	49.025	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	98.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121124\
Data File : VX044228.D
Acq On : 11 Dec 2024 14:51
Operator : JC/MD
Sample : VX1211WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1211WBL01

Quant Time: Dec 12 04:52:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

