

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121125\
 Data File : VX048831.D
 Acq On : 11 Dec 2025 17:00
 Operator : JC/MD
 Sample : Q3835-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OWBR-01-170-121025

Quant Time: Dec 11 23:53:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) Pentafluorobenzene		5.544	168	176493	50.000	ug/l	0.00
34) 1,4-Difluorobenzene		6.745	114	295769	50.000	ug/l	0.00
63) Chlorobenzene-d5		10.037	117	264330	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4		12.006	152	123284	50.000	ug/l	0.00
System Monitoring Compounds							
33) 1,2-Dichloroethane-d4		5.940	65	123273	52.039	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.080%	
35) Dibromofluoromethane		5.373	113	97951	48.098	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.200%	
50) Toluene-d8		8.634	98	354910	49.718	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.440%	
62) 4-Bromofluorobenzene		11.061	95	120279	48.862	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.720%	
Target Compounds							
						Qvalue	
27) cis-1,2-Dichloroethene		4.483	96	3471	1.471	ug/l	92
44) Trichloroethene		7.110	130	15008	6.956	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121125\
Data File : VX048831.D
Acq On : 11 Dec 2025 17:00
Operator : JC/MD
Sample : Q3835-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OWBR-01-170-121025

Quant Time: Dec 11 23:53:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

