

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121125\
 Data File : VX048834.D
 Acq On : 11 Dec 2025 18:02
 Operator : JC/MD
 Sample : Q3828-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BR-02-133-120925

Quant Time: Dec 11 23:54:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	192834	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.750	114	321114	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.036	117	287581	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	130094	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	138177	53.388	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.780%
35) Dibromofluoromethane	5.373	113	107973	48.835	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.660%
50) Toluene-d8	8.634	98	388530	50.131	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.260%
62) 4-Bromofluorobenzene	11.061	95	124801	46.698	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.400%
Target Compounds						
						Qvalue
44) Trichloroethene	7.116	130	12613	5.385	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121125\
Data File : VX048834.D
Acq On : 11 Dec 2025 18:02
Operator : JC/MD
Sample : Q3828-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BR-02-133-120925

Quant Time: Dec 11 23:54:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

