

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121125\
 Data File : VX048835.D
 Acq On : 11 Dec 2025 18:23
 Operator : JC/MD
 Sample : Q3828-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OWBR-06-219-120925

Quant Time: Dec 11 23:54:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	182215	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	305374	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	274085	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	130417	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	133273	54.494	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.980%
35) Dibromofluoromethane	5.373	113	103640	49.291	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.580%
50) Toluene-d8	8.634	98	363009	49.253	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.500%
62) 4-Bromofluorobenzene	11.061	95	122044	48.020	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.040%
Target Compounds						
44) Trichloroethene	7.110	130	6815	3.059	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121125\
Data File : VX048835.D
Acq On : 11 Dec 2025 18:23
Operator : JC/MD
Sample : Q3828-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OWBR-06-219-120925

Quant Time: Dec 11 23:54:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

