

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121125\
 Data File : VX048837.D
 Acq On : 11 Dec 2025 19:05
 Operator : JC/MD
 Sample : Q3828-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OWBR-03-138-120925

Quant Time: Dec 11 23:55:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	222135	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	364309	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	323691	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	138930	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	145593	48.833	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.660%
35) Dibromofluoromethane	5.373	113	118953	47.422	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.840%
50) Toluene-d8	8.635	98	438075	49.822	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.640%
62) 4-Bromofluorobenzene	11.061	95	137953	45.499	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.000%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.337	101	7432	2.941	ug/l	99
27) cis-1,2-Dichloroethene	4.489	96	5863	1.974	ug/l	99
40) Benzene	6.025	78	4659	0.457	ug/l #	87
44) Trichloroethene	7.117	130	9342	3.515	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121125\
Data File : VX048837.D
Acq On : 11 Dec 2025 19:05
Operator : JC/MD
Sample : Q3828-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OWBR-03-138-120925

Quant Time: Dec 11 23:55:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

