

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121218\
 Data File : VX006446.D
 Acq On : 12 Dec 2018 15:46
 Operator : JC/MD
 Sample : J6358-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW308S-20181210

Quant Time: Dec 13 09:57:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	729568	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1127226	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	1016874	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	491551	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	369767	49.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.30%	
35) Dibromofluoromethane	5.50	113	338096	46.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.66%	
50) Toluene-d8	8.71	98	1312392	49.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.88%	
62) 4-Bromofluorobenzene	11.13	95	484636	48.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.90%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121218\
Data File : VX006446.D
Acq On : 12 Dec 2018 15:46
Operator : JC/MD
Sample : J6358-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW308S-20181210

Quant Time: Dec 13 09:57:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

