

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121218\
 Data File : VX006455.D
 Acq On : 12 Dec 2018 19:46
 Operator : JC/MD
 Sample : J6376-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TEST-PIT-1-WATER

Quant Time: Dec 13 10:01:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	671640	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1062964	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	953277	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	445845	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	358265	52.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.50%	
35) Dibromofluoromethane	5.51	113	316811	46.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.06%	
50) Toluene-d8	8.71	98	1228408	49.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.14%	
62) 4-Bromofluorobenzene	11.13	95	454621	48.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.38%	
Target Compounds						
16) Acetone	2.45	43	16590	5.156	ug/l	95
25) 2-Butanone	4.70	43	10435	2.273	ug/l #	77
29) Tetrahydrofuran	5.16	42	35206	11.034	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121218\
Data File : VX006455.D
Acq On : 12 Dec 2018 19:46
Operator : JC/MD
Sample : J6376-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TEST-PIT-1-WATER

Quant Time: Dec 13 10:01:54 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

