

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121218\
 Data File : VX006467.D
 Acq On : 13 Dec 2018 01:59
 Operator : JC/MD
 Sample : J6193-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CL-01-121018-B

Quant Time: Dec 13 07:34:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	673363	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1048260	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	971614	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	481608	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	351890	51.19	ug/l	0.00
Spiked Amount			Recovery	=	102.38%	
35) Dibromofluoromethane	5.51	113	317250	46.74	ug/l	0.00
Spiked Amount			Recovery	=	93.48%	
50) Toluene-d8	8.71	98	1243190	50.36	ug/l	0.00
Spiked Amount			Recovery	=	100.72%	
62) 4-Bromofluorobenzene	11.13	95	467621	50.27	ug/l	0.00
Spiked Amount			Recovery	=	100.54%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	28153	8.727	ug/l	98
18) Methyl Acetate	2.78	43	11257	1.043	ug/l #	87
20) Methylene Chloride	2.86	84	87671	12.770	ug/l	91
43) Isopropyl Acetate	6.46	43	37970	2.304	ug/l	98
52) Toluene	8.79	92	292413	16.811	ug/l	98
67) Ethyl Benzene	10.25	91	227529	6.633	ug/l	98
68) m/p-Xylenes	10.36	106	383163	27.884	ug/l	99
69) o-Xylene	10.70	106	142742	10.517	ug/l	98
78) n-propylbenzene	11.36	91	96203	2.602	ug/l	100
80) 1,3,5-Trimethylbenzene	11.51	105	163018	5.891	ug/l	99
84) 1,2,4-Trimethylbenzene	11.80	105	586900	20.539	ug/l	99
95) Naphthalene	13.83	128	142995	3.866	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121218\
Data File : VX006467.D
Acq On : 13 Dec 2018 01:59
Operator : JC/MD
Sample : J6193-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CL-01-121018-B

Quant Time: Dec 13 07:34:51 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

