

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121218\
 Data File : VX006471.D
 Acq On : 13 Dec 2018 03:46
 Operator : JC/MD
 Sample : J6189-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HR-05-121118-A

Quant Time: Dec 13 07:41:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	683072	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1067126	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	982635	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	470256	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	358106	51.36	ug/l	0.00
Spiked Amount			Recovery	=	102.72%	
35) Dibromofluoromethane	5.51	113	320961	46.45	ug/l	0.00
Spiked Amount			Recovery	=	92.90%	
50) Toluene-d8	8.71	98	1270661	50.56	ug/l	0.00
Spiked Amount			Recovery	=	101.12%	
62) 4-Bromofluorobenzene	11.13	95	464348	49.03	ug/l	0.00
Spiked Amount			Recovery	=	98.06%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	27474	8.396	ug/l	97
18) Methyl Acetate	2.78	43	13347	1.220	ug/l	100
20) Methylene Chloride	2.86	84	402441	57.785	ug/l	95
43) Isopropyl Acetate	6.46	43	33068	1.971	ug/l	98
84) 1,2,4-Trimethylbenzene	11.80	105	36424	1.305	ug/l	97
95) Naphthalene	13.83	128	75577	2.092	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121218\
Data File : VX006471.D
Acq On : 13 Dec 2018 03:46
Operator : JC/MD
Sample : J6189-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HR-05-121118-A

Quant Time: Dec 13 07:41:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

