

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121218\
 Data File : VX006482.D
 Acq On : 13 Dec 2018 08:39
 Operator : JC/MD
 Sample : J6347-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P001-SW002-01

Quant Time: Dec 13 12:15:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	659855	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1043633	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	948054	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	441469	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	352919	52.39	ug/l	0.00
Spiked Amount			Recovery	=	104.78%	
35) Dibromofluoromethane	5.51	113	311686	46.13	ug/l	0.00
Spiked Amount			Recovery	=	92.26%	
50) Toluene-d8	8.71	98	1226938	49.92	ug/l	0.00
Spiked Amount			Recovery	=	99.84%	
62) 4-Bromofluorobenzene	11.13	95	447034	48.27	ug/l	0.00
Spiked Amount			Recovery	=	96.54%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121218\
Data File : VX006482.D
Acq On : 13 Dec 2018 08:39
Operator : JC/MD
Sample : J6347-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P001-SW002-01

Quant Time: Dec 13 12:15:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. 0.00 min

Lab File: VX006482.D

Acq: 13 Dec 2018 08:39

Instrument :

MSVOA_X

ClientSampleId :


P001-SW002-01

Tot Ion:168 Resp: 659855

Ion Ratio Lower Upper




168 100

99 53.2 37.8 56.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 52.392 ug/l

RT: 6.07 min Scan# 806

Delta R.T. 0.00 min

Lab File: VX006482.D


Acq: 13 Dec 2018 08:39

Tgt Ion: 65 Resp: 352919

Ion Ratio Lower Upper



65 100


67 53.4 0.0 107.2



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.86 min Scan# 936

Delta R.T. 0.00 min

Lab File: VX006482.D

Acq: 13 Dec 2018 08:39

Instrument :

MSVOA_X

ClientSampleId :

P001-SW002-01


Tot Ion:114 Resp: 1043633

Ion Ratio Lower Upper

114 100

63 18.1 0.0 36.6




88 15.7 0.0 29.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 46.127 ug/l

RT: 5.51 min Scan# 714

Delta R.T. 0.01 min

Lab File: VX006482.D

Acq: 13 Dec 2018 08:39


Tgt Ion:113 Resp: 311686

Ion Ratio Lower Upper

113 100

111 102.7 81.1 121.7



192 20.3 18.2 27.2




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 49.919 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. 0.00 min

Lab File: VX006482.D

Acq: 13 Dec 2018 08:39

Instrument :

MSVOA_X

ClientSampleId :





P001-SW002-01

Tot Ion: 98 Resp: 1226938

Ion Ratio Lower Upper

98 100

100 64.2 52.2 78.2



#62

4-Bromofluorobenzene

Concen: 48.266 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. -0.01 min

Lab File: VX006482.D

Acq: 13 Dec 2018 08:39




Tgt Ion: 95 Resp: 447034


Ion Ratio Lower Upper

95 100

174 86.6 0.0 187.0

176 83.9 0.0 181.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.12 min Scan# 1470
Delta R.T. -0.00 min
Lab File: VX006482.D
Acq: 13 Dec 2018 08:39

Instrument :
MSVOA_X
ClientSampleId :
P001-SW002-01

Tot Ion:117 Resp: 948054
Ion Ratio Lower Upper
117 100
82 50.6 39.6 59.4
119 32.0 26.3 39.5




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. 0.00 min
Lab File: VX006482.D
Acq: 13 Dec 2018 08:39

Tgt Ion:152 Resp: 441469
Ion Ratio Lower Upper
152 100
115 57.8 39.0 116.9
150 159.9 0.0 345.8



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

