

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121219\
 Data File : VX013974.D
 Acq On : 12 Dec 2019 14:49
 Operator : JC/SP
 Sample : K6252-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB03-20191210

Quant Time: Dec 13 03:36:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	572071	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	870069	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	768638	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	321172	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	318321	47.94	ug/l	0.00
Spiked Amount			Recovery	=	95.88%	
35) Dibromofluoromethane	5.49	113	255406	47.48	ug/l	0.00
Spiked Amount			Recovery	=	94.96%	
50) Toluene-d8	8.71	98	1025264	48.92	ug/l	0.00
Spiked Amount			Recovery	=	97.84%	
62) 4-Bromofluorobenzene	11.14	95	334029	44.20	ug/l	0.00
Spiked Amount			Recovery	=	88.40%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121219\
Data File : VX013974.D
Acq On : 12 Dec 2019 14:49
Operator : JC/SP
Sample : K6252-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB03-20191210

Quant Time: Dec 13 03:36:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

