

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121219\
 Data File : VX013984.D
 Acq On : 12 Dec 2019 18:42
 Operator : JC/SP
 Sample : K6252-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D5-20191209

Quant Time: Dec 13 04:25:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	551145	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	851347	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	753888	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	316837	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	309820	48.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.86%	
35) Dibromofluoromethane	5.48	113	252742	48.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.04%	
50) Toluene-d8	8.71	98	1007142	49.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.24%	
62) 4-Bromofluorobenzene	11.14	95	329005	44.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.98%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.19	85	6864	1.119	ug/l	91
9) 1,1,2-Trichlorotrifluoroet	2.37	101	55204	10.506	ug/l	98
12) 1,1-Dichloroethene	2.36	96	11916	2.272	ug/l	96
27) cis-1,2-Dichloroethene	4.59	96	9494	1.447	ug/l	88
38) Carbon Tetrachloride	5.76	117	6433	0.903	ug/l #	76
44) Trichloroethene	7.20	130	701622	108.615	ug/l	100
49) 1,4-Dioxane	7.75	88	2538	16.350	ug/l #	62
52) Toluene	8.78	92	7670	0.525	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121219\
Data File : VX013984.D
Acq On : 12 Dec 2019 18:42
Operator : JC/SP
Sample : K6252-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D5-20191209

Quant Time: Dec 13 04:25:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

