

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121219\
 Data File : VX014002.D
 Acq On : 13 Dec 2019 04:17
 Operator : JC/SP
 Sample : K6258-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-NEAR-THE-CURB-(3A)

Quant Time: Dec 13 05:01:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X121219W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Dec 13 01:35:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.00	128	92253	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.85	114	491367	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.11	117	425521	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.05	65	197372	29.87	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	99.57%
60) 4-Bromofluorobenzene	11.13	95	186096	27.08	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	90.27%
63) Toluene-d8	8.71	98	581865	30.59	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	101.97%

Target Compounds					Ovalue
15) Acetone	2.44	58	5761	6.273 ug/l	# 62

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121219\
Data File : VX014002.D
Acq On : 13 Dec 2019 04:17
Operator : JC/SP
Sample : K6258-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-NEAR-THE-CURB-(3A)

Quant Time: Dec 13 05:01:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X121219W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Dec 13 01:35:34 2019
Response via : Initial Calibration

