

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121219\
 Data File : VX013975.D
 Acq On : 12 Dec 2019 15:13
 Operator : JC/SP
 Sample : K6252-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB02-20191209

Quant Time: Dec 13 03:39:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	558471	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	857008	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	756762	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	317787	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	315479	48.67	ug/l	0.00
Spiked Amount			Recovery	=	97.34%	
35) Dibromofluoromethane	5.49	113	251479	47.47	ug/l	0.00
Spiked Amount			Recovery	=	94.94%	
50) Toluene-d8	8.71	98	1016743	49.26	ug/l	0.00
Spiked Amount			Recovery	=	98.52%	
62) 4-Bromofluorobenzene	11.14	95	331273	44.50	ug/l	0.00
Spiked Amount			Recovery	=	89.00%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121219\
Data File : VX013975.D
Acq On : 12 Dec 2019 15:13
Operator : JC/SP
Sample : K6252-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB02-20191209

Quant Time: Dec 13 03:39:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.65 min Scan# 781

Delta R.T. 0.00 min

Lab File: VX013975.D

Acq: 12 Dec 2019 15:13

Instrument :

MSVOA_X

ClientSampleId :


FB02-20191209

Tot Ion:168 Resp: 558471

Ion Ratio Lower Upper

168 100

99 49.5 39.8 59.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.65

200000

150000



100000

50000

0

Time-->

5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 48.666 ug/l

RT: 6.05 min Scan# 847

Delta R.T. 0.00 min

Lab File: VX013975.D


Acq: 12 Dec 2019 15:13

Tgt Ion: 65 Resp: 315479

Ion Ratio Lower Upper

65 100

67 53.6 0.0 104.4



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.05


100000


50000

0

Time-->

5.90 6.00 6.10 6.20





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.85 min Scan# 978

Delta R.T. 0.00 min

Lab File: VX013975.D

Acq: 12 Dec 2019 15:13

Instrument :

MSVOA_X

ClientSampleId :

FB02-20191209


Tot Ion:114 Resp: 857008

Ion Ratio Lower Upper

114 100

63 20.4 0.0 39.6




88 15.6 0.0 31.4



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 47.466 ug/l

RT: 5.49 min Scan# 755

Delta R.T. 0.00 min

Lab File: VX013975.D

Acq: 12 Dec 2019 15:13


Tgt Ion:113 Resp: 251479

Ion Ratio Lower Upper

113 100

111 103.5 83.6 125.4



192 23.8 19.2 28.8




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 49.256 ug/l

RT: 8.71 min Scan# 1283

Delta R.T. 0.00 min

Lab File: VX013975.D

Acq: 12 Dec 2019 15:13

Instrument :

MSVOA_X

ClientSampleId :





FB02-20191209

Tot Ion: 98 Resp: 1016743

Ion Ratio Lower Upper

98 100

100 65.8 53.1 79.7



#62

4-Bromofluorobenzene

Concen: 44.504 ug/l

RT: 11.14 min Scan# 1681

Delta R.T. 0.00 min

Lab File: VX013975.D

Acq: 12 Dec 2019 15:13




Tgt Ion: 95 Resp: 331273


Ion Ratio Lower Upper

95 100

174 88.4 0.0 180.0

176 85.5 0.0 173.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1513
Delta R.T. 0.00 min
Lab File: VX013975.D
Acq: 12 Dec 2019 15:13

Instrument :
MSVOA_X
ClientSampleId :
FB02-20191209

Tot Ion:117 Resp: 756762
Ion Ratio Lower Upper
117 100
82 52.0 41.8 62.8
119 31.5 25.8 38.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.07 min Scan# 1835
Delta R.T. 0.00 min
Lab File: VX013975.D
Acq: 12 Dec 2019 15:13

Tgt Ion:152 Resp: 317787
Ion Ratio Lower Upper
152 100
115 55.9 38.4 115.1
150 156.3 0.0 346.8

