

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121220\
 Data File : VX019933.D
 Acq On : 12 Dec 2020 15:38
 Operator : JC/MD
 Sample : L4996-17
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D1-20201207

Quant Time: Dec 14 06:05:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	311928	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	514624	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	469575	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	206887	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		6.03	65	197770	48.24	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.48%	
35) Dibromofluoromethane		5.46	113	158878	47.77	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.54%	
50) Toluene-d8		8.70	98	628873	49.53	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.06%	
62) 4-Bromofluorobenzene		11.12	95	218388	45.19	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.38%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
30) Chloroform	5.17	83	1499	0.248	ug/l	96
44) Trichloroethene	7.19	130	19785	5.308	ug/l	100
64) Tetrachloroethene	9.32	164	2207	0.701	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121220\
Data File : VX019933.D
Acq On : 12 Dec 2020 15:38
Operator : JC/MD
Sample : L4996-17
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20201207

Quant Time: Dec 14 06:05:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

