

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121220\
 Data File : VX019934.D
 Acq On : 12 Dec 2020 16:01
 Operator : JC/MD
 Sample : PB133524TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB133524TB

Quant Time: Dec 14 08:31:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	322128	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	531429	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	493426	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	214532	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	205888	48.63	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.26%
35) Dibromofluoromethane	5.46	113	166107	48.36	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.72%
50) Toluene-d8	8.70	98	656016	50.03	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.06%
62) 4-Bromofluorobenzene	11.12	95	230710	46.23	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.46%

Target Compounds

					Qvalue
16) Acetone	2.42	43	12244	7.694	ug/l 91
20) Methylene Chloride	2.83	84	9255	1.158	ug/l 97
43) Isopropyl Acetate	6.41	43	27522	3.491	ug/l 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121220\
Data File : VX019934.D
Acq On : 12 Dec 2020 16:01
Operator : JC/MD
Sample : PB133524TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB133524TB

Quant Time: Dec 14 08:31:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

