

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121220\
 Data File : VX019938.D
 Acq On : 12 Dec 2020 17:34
 Operator : JC/MD
 Sample : PB133524ZHE#02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB133524ZHE#02

Quant Time: Dec 14 08:37:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	301816	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	507403	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	476578	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	200987	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	197530	49.80	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	99.60%		
35) Dibromofluoromethane	5.46	113	158905	48.46	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	96.92%		
50) Toluene-d8	8.70	98	628178	50.18	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	100.36%		
62) 4-Bromofluorobenzene	11.12	95	219042	45.97	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	91.94%		

Target Compounds

					Qvalue
16) Acetone	2.42	43	32965	22.109 ug/l	98
20) Methylene Chloride	2.83	84	78266	22.202 ug/l	98
43) Isopropyl Acetate	6.41	43	24175	3.212 ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121220\
Data File : VX019938.D
Acq On : 12 Dec 2020 17:34
Operator : JC/MD
Sample : PB133524ZHE#02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB133524ZHE#02

Quant Time: Dec 14 08:37:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

