

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121220\
 Data File : VX019942.D
 Acq On : 12 Dec 2020 19:07
 Operator : JC/MD
 Sample : PB133524ZHE#06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB133524ZHE#06

Quant Time: Dec 14 08:41:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	296517	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	484166	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	461458	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	186948	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	187688	48.16	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.32%
35) Dibromofluoromethane	5.46	113	153361	49.01	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.02%
50) Toluene-d8	8.70	98	607769	50.88	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.76%
62) 4-Bromofluorobenzene	11.12	95	207491	45.63	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.26%

Target Compounds

					Qvalue
16) Acetone	2.42	43	33097	22.595	ug/l 100
20) Methylene Chloride	2.83	84	77180	22.291	ug/l 97
43) Isopropyl Acetate	6.41	43	23752	3.307	ug/l 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121220\
Data File : VX019942.D
Acq On : 12 Dec 2020 19:07
Operator : JC/MD
Sample : PB133524ZHE#06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB133524ZHE#06

Quant Time: Dec 14 08:41:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

