

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121220\
 Data File : VX019943.D
 Acq On : 12 Dec 2020 19:31
 Operator : JC/MD
 Sample : PB133524ZHE#07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB133524ZHE#07

Quant Time: Dec 14 08:42:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	300402	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	501923	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	472846	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	195427	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	195679	49.56	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.12%
35) Dibromofluoromethane	5.46	113	157706	48.62	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.24%
50) Toluene-d8	8.70	98	630453	50.91	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.82%
62) 4-Bromofluorobenzene	11.12	95	214699	45.55	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.10%

Target Compounds

					Qvalue
16) Acetone	2.42	43	11753	7.920	ug/l 96
20) Methylene Chloride	2.83	84	8722	1.186	ug/l 96
43) Isopropyl Acetate	6.42	43	27712	3.722	ug/l 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121220\
Data File : VX019943.D
Acq On : 12 Dec 2020 19:31
Operator : JC/MD
Sample : PB133524ZHE#07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB133524ZHE#07

Quant Time: Dec 14 08:42:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

