

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121220\
 Data File : VX019944.D
 Acq On : 12 Dec 2020 19:54
 Operator : JC/MD
 Sample : PB133524ZHE#08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB133524ZHE#08

Quant Time: Dec 14 08:43:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	301352	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	505179	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	479430	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	203720	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	196652	49.65	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	99.30%		
35) Dibromofluoromethane	5.46	113	159666	48.90	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	97.80%		
50) Toluene-d8	8.70	98	636585	51.07	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	102.14%		
62) 4-Bromofluorobenzene	11.12	95	220647	46.51	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	93.02%		

Target Compounds					Qvalue
16) Acetone	2.42	43	12347	8.294 ug/l	99
20) Methylene Chloride	2.84	84	8830	1.210 ug/l	93
43) Isopropyl Acetate	6.41	43	28289	3.775 ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121220\
Data File : VX019944.D
Acq On : 12 Dec 2020 19:54
Operator : JC/MD
Sample : PB133524ZHE#08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB133524ZHE#08

Quant Time: Dec 14 08:43:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

