

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121220\
 Data File : VX019946.D
 Acq On : 12 Dec 2020 20:40
 Operator : JC/MD
 Sample : L4984-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TR3-C-TOP

Quant Time: Dec 14 08:46:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	301805	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	504481	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	480450	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	206562	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		6.03	65	197106	49.69	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.38%	
35) Dibromofluoromethane		5.46	113	158388	48.58	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.16%	
50) Toluene-d8		8.70	98	629480	50.57	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.14%	
62) 4-Bromofluorobenzene		11.12	95	220961	46.64	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.28%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.42	43	26244	17.602	ug/l	98
20) Methylene Chloride	2.83	84	35018	9.125	ug/l	96
43) Isopropyl Acetate	6.42	43	19752	2.640	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121220\
Data File : VX019946.D
Acq On : 12 Dec 2020 20:40
Operator : JC/MD
Sample : L4984-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TR3-C-TOP

Quant Time: Dec 14 08:46:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

