

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121220\
 Data File : VX019949.D
 Acq On : 12 Dec 2020 21:50
 Operator : JC/MD
 Sample : L4984-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TR3-D-BOTTOM

Quant Time: Dec 14 08:49:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	307333	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	509128	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	476972	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	203413	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	198777	49.21	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	98.42%		
35) Dibromofluoromethane	5.46	113	159097	48.35	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	96.70%		
50) Toluene-d8	8.70	98	629471	50.11	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	100.22%		
62) 4-Bromofluorobenzene	11.12	95	218027	45.60	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	91.20%		


Target Compounds					Qvalue
16) Acetone	2.42	43	26628	17.539 ug/l	99
20) Methylene Chloride	2.83	84	33110	8.368 ug/l	98
43) Isopropyl Acetate	6.41	43	23118	3.061 ug/l	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121220\
Data File : VX019949.D
Acq On : 12 Dec 2020 21:50
Operator : JC/MD
Sample : L4984-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TR3-D-BOTTOM

Quant Time: Dec 14 08:49:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.63 min Scan# 745

Delta R.T. 0.01 min

Lab File: VX019949.D

Acq: 12 Dec 2020 21:50

Instrument :

MSVOA_X

ClientSampleId :


TR3-D-BOTTOM

Tot Ion:168 Resp: 307333

Ion Ratio Lower Upper




168 100

99 55.1 45.8 68.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#16

Acetone

Concen: 17.539 ug/l

RT: 2.42 min Scan# 218

Delta R.T. -0.00 min

Lab File: VX019949.D


Acq: 12 Dec 2020 21:50

Tgt Ion: 43 Resp: 26628

Ion Ratio Lower Upper



43 100


58 35.3 27.8 41.8



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0









#20
Methvlene Chloride
Concen: 8.368 ug/l
RT: 2.83 min Scan# 286
Delta R.T. 0.00 min
Lab File: VX019949.D
Acq: 12 Dec 2020 21:50




Instrument :
MSVOA_X
ClientSampleId :
TR3-D-BOTTOM


Tot Ion:	84	Resp:	33110
Ion	Ratio	Lower	Upper
84	100		
49	112.5	88.6	132.8
51	35.8	27.0	40.6
86	62.1	50.3	75.5



#33
1,2-Dichloroethane-d4
Concen: 49.211 ug/l
RT: 6.03 min Scan# 811
Delta R.T. 0.01 min
Lab File: VX019949.D
Acq: 12 Dec 2020 21:50

Tgt Ion:	65	Resp:	198777
Ion	Ratio	Lower	Upper
65	100		
67	53.2	0.0	105.4









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.83 min Scan# 942
 Delta R.T. 0.01 min
 Lab File: VX019949.D
 Acq: 12 Dec 2020 21:50

Instrument :
 MSVOA_X
 ClientSampleId :
 TR3-D-BOTTOM

Tot Ion	Ratio	Lower	Upper
114	100		
63	21.0	0.0	40.8
88	16.5	0.0	32.8




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VX0
 Ion 87.90 (87.60 to 88.60): VX0






#35
 Dibromofluoromethane
 Concen: 48.352 ug/l
 RT: 5.46 min Scan# 718
 Delta R.T. 0.01 min
 Lab File: VX019949.D
 Acq: 12 Dec 2020 21:50

Tgt Ion	Ratio	Lower	Upper
113	100		
111	102.5	81.9	122.9
192	20.4	16.7	25.1



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#43

Isopropyl Acetate

Concen: 3.061 ug/l

RT: 6.41 min Scan# 873

Delta R.T. 0.01 min

Lab File: VX019949.D

Acq: 12 Dec 2020 21:50

Instrument :

MSVOA_X

ClientSampleId :

TR3-D-BOTTOM





Tot Ion: 43 Resp: 23118

Ion Ratio Lower Upper

43 100

61 22.0 18.2 27.4

87 14.4 12.5 18.7



#50

Toluene-d8

Concen: 50.109 ug/l

RT: 8.70 min Scan# 1248

Delta R.T. -0.00 min

Lab File: VX019949.D




Acq: 12 Dec 2020 21:50


Tgt Ion: 98 Resp: 629471

Ion Ratio Lower Upper

98 100

100 65.6 51.9 77.9





#62

4-Bromofluorobenzene

Concen: 45.600 ug/l

RT: 11.12 min Scan# 1646

Delta R.T. -0.00 min

Lab File: VX019949.D

Acq: 12 Dec 2020 21:50

Instrument :

MSVOA_X

ClientSampleId :

TR3-D-BOTTOM


Tot Ion: 95 Resp: 218027

Ion Ratio Lower Upper

95 100

174 76.2 0.0 154.6




176 73.3 0.0 155.8



Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): VX0

Ion 175.80 (175.50 to 176.50): VX0



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.10 min Scan# 1478

Delta R.T. 0.01 min

Lab File: VX019949.D

Acq: 12 Dec 2020 21:50


Tgt Ion: 117 Resp: 476972

Ion Ratio Lower Upper

117 100

82 56.3 47.4 71.2



119 32.1 26.6 39.8




Abundance Ion 116.90 (116.60 to 117.60): VX0

Ion 82.00 (81.70 to 82.70): VX0

Ion 118.90 (118.60 to 119.60): VX0





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.06 min Scan# 1800

Delta R.T. -0.00 min

Lab File: VX019949.D

Acq: 12 Dec 2020 21:50

Instrument :

MSVOA_X

ClientSampleId :

TR3-D-BOTTOM


Tot Ion:152 Resp: 203413

Ion Ratio Lower Upper

152 100

115 60.4 41.1 123.3

150 156.3 0.0 349.0



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

