

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121222\
 Data File : VX033370.D
 Acq On : 12 Dec 2022 13:24
 Operator : JC/MD
 Sample : VX1212WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1212WBL01

Quant Time: Dec 13 07:30:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X121222W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Dec 12 12:37:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	12813	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	87498	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	92776	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	35197	31.278	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	104.267%
60) 4-Bromofluorobenzene	11.079	95	48163	28.029	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	93.433%
63) Toluene-d8	8.647	98	92634	30.016	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.067%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121222\
Data File : VX033370.D
Acq On : 12 Dec 2022 13:24
Operator : JC/MD
Sample : VX1212WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1212WBL01

Quant Time: Dec 13 07:30:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X121222W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Dec 12 12:37:35 2022
Response via : Initial Calibration

