

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121223\
 Data File : VX039155.D
 Acq On : 12 Dec 2023 10:38
 Operator : JC/MD
 Sample : VX1212MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1212MBL01

Quant Time: Dec 13 04:12:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	104074	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	213120	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	214969	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	103607	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	89886	53.600	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.200%
35) Dibromofluoromethane	5.379	113	68253	44.343	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.680%
50) Toluene-d8	8.647	98	266805	45.305	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.620%#
62) 4-Bromofluorobenzene	11.079	95	117360	49.783	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121223\
Data File : VX039155.D
Acq On : 12 Dec 2023 10:38
Operator : JC/MD
Sample : VX1212MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1212MBL01

Quant Time: Dec 13 04:12:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

