

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121223\
 Data File : VX039156.D
 Acq On : 12 Dec 2023 11:01
 Operator : JC/MD
 Sample : VX1212WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1212WBL01

Quant Time: Dec 13 04:13:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	144487	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	293503	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	294789	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	142596	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	120773	51.874	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.740%	
35) Dibromofluoromethane	5.385	113	94944	44.790	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 89.580%	
50) Toluene-d8	8.647	98	367931	45.366	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 90.740%#	
62) 4-Bromofluorobenzene	11.079	95	162731	50.124	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.240%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121223\
Data File : VX039156.D
Acq On : 12 Dec 2023 11:01
Operator : JC/MD
Sample : VX1212WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1212WBL01

Quant Time: Dec 13 04:13:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

