

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121223\
 Data File : VX039160.D
 Acq On : 12 Dec 2023 13:08
 Operator : JC/MD
 Sample : 05639-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 E8965C-CC-001

Quant Time: Dec 13 04:15:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	121089	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	244706	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	245001	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	118095	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	101226	51.880	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 103.760%			
35) Dibromofluoromethane	5.385	113	78344	44.329	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 88.660%			
50) Toluene-d8	8.647	98	309915	45.833	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 91.660%#			
62) 4-Bromofluorobenzene	11.079	95	136728	50.512	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 101.020%			
Target Compounds						
11) Tert butyl alcohol	2.946	59	7644	14.086	ug/l #	92
16) Acetone	2.373	43	31236	30.073	ug/l	94
20) Methylene Chloride	2.788	84	10088	5.388	ug/l	98
25) 2-Butanone	4.556	43	33751	20.038	ug/l	97
43) Isopropyl Acetate	6.342	43	14617	2.772	ug/l	99
59) 2-Hexanone	9.433	43	13647	5.052	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121223\
Data File : VX039160.D
Acq On : 12 Dec 2023 13:08
Operator : JC/MD
Sample : 05639-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
E8965C-CC-001

Quant Time: Dec 13 04:15:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

