

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121223\
 Data File : VX039162.D
 Acq On : 12 Dec 2023 13:53
 Operator : JC/MD
 Sample : 05691-03DL 40X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 A3909DL

Quant Time: Dec 13 04:15:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	112011	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	232084	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	239581	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	119860	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	97376	53.951	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.900%
35) Dibromofluoromethane	5.385	113	75203	44.866	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.740%
50) Toluene-d8	8.647	98	291314	45.425	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.840%#
62) 4-Bromofluorobenzene	11.079	95	133333	51.937	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.880%
Target Compounds						
16) Acetone	2.374	43	94659	98.520	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121223\
Data File : VX039162.D
Acq On : 12 Dec 2023 13:53
Operator : JC/MD
Sample : 05691-03DL 40X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
A3909DL

Quant Time: Dec 13 04:15:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

