

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121223\
 Data File : VX039165.D
 Acq On : 12 Dec 2023 15:02
 Operator : JC/MD
 Sample : 05729-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Dec 13 04:16:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	108775	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	227333	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	230727	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	118927	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	94455	53.890	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.780%	
35) Dibromofluoromethane	5.385	113	73901	45.011	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 90.020%	
50) Toluene-d8	8.647	98	283842	45.185	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 90.360%#	
62) 4-Bromofluorobenzene	11.079	95	130934	52.068	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.140%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121223\
Data File : VX039165.D
Acq On : 12 Dec 2023 15:02
Operator : JC/MD
Sample : 05729-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Dec 13 04:16:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

