

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121223\
 Data File : VX039166.D
 Acq On : 12 Dec 2023 15:25
 Operator : JC/MD
 Sample : 05729-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Dec 13 04:17:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	109760	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	225853	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	232985	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	117409	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	95118	53.781	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.560%	
35) Dibromofluoromethane	5.385	113	74651	45.765	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 91.540%	
50) Toluene-d8	8.647	98	283694	45.457	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 90.920%#	
62) 4-Bromofluorobenzene	11.079	95	130050	52.056	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.120%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121223\
Data File : VX039166.D
Acq On : 12 Dec 2023 15:25
Operator : JC/MD
Sample : 05729-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Dec 13 04:17:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

