

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121223\
 Data File : VX039167.D
 Acq On : 12 Dec 2023 15:48
 Operator : JC/MD
 Sample : 05719-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TB-20231205

Quant Time: Dec 13 04:17:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	110411	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	227744	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	235171	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	119739	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	96371	54.168	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.340%
35) Dibromofluoromethane	5.385	113	73710	44.813	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.620%
50) Toluene-d8	8.647	98	288076	45.776	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.560%#
62) 4-Bromofluorobenzene	11.079	95	133597	53.032	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.060%
Target Compounds						
16) Acetone	2.380	43	2952	3.117	ug/l	Qvalue # 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121223\
Data File : VX039167.D
Acq On : 12 Dec 2023 15:48
Operator : JC/MD
Sample : 05719-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TB-20231205

Quant Time: Dec 13 04:17:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

