

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121223\
 Data File : VX039175.D
 Acq On : 12 Dec 2023 18:51
 Operator : JC/MD
 Sample : 05718-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-SOIL-20231206

Quant Time: Dec 13 04:20:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	108406	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	221123	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	229635	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	117217	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	95304	54.559	ug/l	0.00
Spiked Amount 50.000	Range	78 - 117	Recovery	=	109.120%	
35) Dibromofluoromethane	5.385	113	72841	45.611	ug/l	0.00
Spiked Amount 50.000	Range	75 - 124	Recovery	=	91.220%	
50) Toluene-d8	8.647	98	282350	46.210	ug/l	0.00
Spiked Amount 50.000	Range	92 - 112	Recovery	=	92.420%	
62) 4-Bromofluorobenzene	11.079	95	130579	53.386	ug/l	0.00
Spiked Amount 50.000	Range	83 - 123	Recovery	=	106.780%	
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	57187	61.499	ug/l	98
18) Methyl Acetate	2.697	43	7488	1.695	ug/l	97
20) Methylene Chloride	2.788	84	7157	4.270	ug/l	96
43) Isopropyl Acetate	6.336	43	115831	24.305	ug/l	100
59) 2-Hexanone	9.372	43	99958	40.948	ug/l #	26

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121223\
Data File : VX039175.D
Acq On : 12 Dec 2023 18:51
Operator : JC/MD
Sample : 05718-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-SOIL-20231206

Quant Time: Dec 13 04:20:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

