

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121224\
 Data File : VX044252.D
 Acq On : 12 Dec 2024 15:09
 Operator : JC/MD
 Sample : P5261-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Dec 13 01:41:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	118608	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	225924	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	196920	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82721	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	104813	50.401	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.800%	
35) Dibromofluoromethane	5.379	113	75423	48.202	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 96.400%	
50) Toluene-d8	8.647	98	275060	52.105	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.220%	
62) 4-Bromofluorobenzene	11.079	95	91071	49.372	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 98.740%	

Target Compounds			Qvalue			
3) Chloromethane	1.294	50	256	0.146	ug/l #	42
4) Vinyl Chloride	1.624	62	209	0.114	ug/l #	43
5) Bromomethane	1.599	94	563	0.434	ug/l #	34
6) Chloroethane	1.691	64	163	0.143	ug/l #	41
7) Trichlorofluoromethane	1.867	101	89	0.025	ug/l #	18
8) Diethyl Ether	2.123	74	34	0.028	ug/l #	1
9) 1,1,2-Trichlorotrifluo...	2.319	101	2741	1.894	ug/l	96
10) Methyl Iodide	2.160	142	21	0.011	ug/l #	34
11) Tert butyl alcohol	2.947	59	6039	20.605	ug/l #	72
12) 1,1-Dichloroethene	2.306	96	566	0.426	ug/l #	79
13) Acrolein	2.203	56	42	0.099	ug/l #	14
14) Allyl chloride	2.666	41	90	0.041	ug/l #	1
15) Acrylonitrile	2.953	53	45	0.055	ug/l #	17
16) Acetone	2.380	43	926	1.092	ug/l #	79
17) Carbon Disulfide	2.501	76	192	0.076	ug/l #	75
18) Methyl Acetate	2.703	43	436	0.199	ug/l #	54
19) Methyl tert-butyl Ether	3.123	73	55	0.011	ug/l #	56
20) Methylene Chloride	2.782	84	482	0.303	ug/l #	64
21) trans-1,2-Dichloroethene	3.081	96	562	0.398	ug/l #	33
23) Vinyl Acetate	3.727	43	20	0.005	ug/l #	73
24) 1,1-Dichloroethane	3.617	63	161	0.058	ug/l #	82
25) 2-Butanone	4.428	43	25	0.021	ug/l #	50
27) cis-1,2-Dichloroethene	4.483	96	2700	1.505	ug/l	80
29) Tetrahydrofuran	5.013	42	19	0.025	ug/l #	32
30) Chloroform	5.092	83	539	0.173	ug/l #	61
31) Cyclohexane	5.550	56	3527	1.571	ug/l #	46
32) 1,1,1-Trichloroethane	5.562	97	19	0.007	ug/l #	1
36) 1,1-Dichloropropene	5.538	75	11353	5.613	ug/l #	46
37) Ethyl Acetate	4.782	43	50	0.020	ug/l #	71
38) Carbon Tetrachloride	5.544	117	13868	6.369	ug/l #	18
39) Methylcyclohexane	7.117	83	5657	2.340	ug/l #	1
40) Benzene	6.044	78	98	0.016	ug/l #	51
41) Methacrylonitrile	4.910	41	75	0.056	ug/l #	1
42) 1,2-Dichloroethane	6.098	62	315	0.125	ug/l #	78
43) Isopropyl Acetate	6.361	43	19	0.005	ug/l #	63
44) Trichloroethene	7.123	130	395453	258.228	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121224\
 Data File : VX044252.D
 Acq On : 12 Dec 2024 15:09
 Operator : JC/MD
 Sample : P5261-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Dec 13 01:41:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
47) Bromodichloromethane	7.836	83	47	0.022	ug/l #	26
48) Methyl methacrylate	7.665	41	31	0.016	ug/l #	1
49) 1,4-Dioxane	7.683	88	56	1.781	ug/l #	32
51) 4-Methyl-2-Pentanone	8.549	43	39	0.016	ug/l #	38
52) Toluene	8.714	92	272	0.071	ug/l	82
53) t-1,3-Dichloropropene	9.000	75	22	1.006	ug/l #	41
55) 1,1,2-Trichloroethane	9.165	97	251	0.164	ug/l #	49
60) Dibromochloromethane	9.281	129	2297	2.674	ug/l #	11
64) Tetrachloroethene	9.275	164	2054	1.572	ug/l #	81
65) Chlorobenzene	10.085	112	86	0.020	ug/l #	43
67) Ethyl Benzene	10.201	91	335	0.046	ug/l #	45
68) m/p-Xylenes	10.305	106	431	0.161	ug/l	87
69) o-Xylene	10.646	106	100	0.037	ug/l #	26
70) Styrene	10.689	104	23	0.005	ug/l #	23
73) Isopropylbenzene	10.976	105	46	0.007	ug/l #	49
74) N-amyl acetate	10.963	43	40	0.014	ug/l #	45
75) 1,1,2,2-Tetrachloroethane	11.079	83	160	0.073	ug/l #	5
76) 1,2,3-Trichloropropane	11.079	75	51556	27.914	ug/l #	35
77) Bromobenzene	11.195	156	19	0.012	ug/l #	1
78) n-propylbenzene	11.311	91	198	0.028	ug/l #	54
79) 2-Chlorotoluene	11.311	91	198	0.044	ug/l #	42
80) 1,3,5-Trimethylbenzene	11.457	105	91	0.017	ug/l	96
81) trans-1,4-Dichloro-2-b...	11.079	75	51556	76.642	ug/l #	4
82) 4-Chlorotoluene	11.463	91	42	0.008	ug/l #	1
83) tert-Butylbenzene	11.725	119	116	0.022	ug/l #	55
84) 1,2,4-Trimethylbenzene	11.762	105	78	0.015	ug/l	86
85) sec-Butylbenzene	11.762	105	78	0.012	ug/l #	57
86) p-Isopropyltoluene	12.012	119	87	0.017	ug/l #	1
87) 1,3-Dichlorobenzene	11.981	146	20	0.007	ug/l #	1
88) 1,4-Dichlorobenzene	12.030	146	312	0.113	ug/l #	1
89) n-Butylbenzene	12.372	91	20	0.005	ug/l #	33
90) Hexachloroethane	12.305	117	40	0.052	ug/l #	19
95) Naphthalene	13.786	128	413	0.069	ug/l #	69
96) 1,2,3-Trichlorobenzene	13.975	180	21	0.013	ug/l #	12


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121224\
Data File : VX044252.D
Acq On : 12 Dec 2024 15:09
Operator : JC/MD
Sample : P5261-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Dec 13 01:41:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09




Instrument :

MSVOA_X

ClientSampleId :







Tgt Ion:168 Resp: 118608
Ion Ratio Lower Upper
168 100
99 67.5 53.4 80.0



#3
Chloromethane
Concen: 0.146 ug/l
RT: 1.294 min Scan# 34
Delta R.T. -0.000 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Tgt Ion: 50 Resp: 256
Ion Ratio Lower Upper
50 100
52 0.0 26.5 39.7#









#4
Vinyl Chloride
Concen: 0.114 ug/l
RT: 1.624 min Scan# 81
Delta R.T. 0.250 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 62 Resp: 209
Ion Ratio Lower Upper
62 100
64 0.0 25.7 38.5#



#5
Bromomethane
Concen: 0.434 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.006 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Tgt Ion: 94 Resp: 563
Ion Ratio Lower Upper
94 100
96 30.3 74.9 112.3#









#6
Chloroethane
Concen: 0.143 ug/l
RT: 1.691 min Scan# 99
Delta R.T. 0.024 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 64 Resp: 163
Ion Ratio Lower Upper
64 100
66 0.0 26.6 39.8#



#7
Trichlorofluoromethane
Concen: 0.025 ug/l
RT: 1.867 min Scan# 128
Delta R.T. -0.000 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09


Tgt Ion:101 Resp: 89
Ion Ratio Lower Upper
101 100
103 0.0 51.8 77.8#








#8
Diethyl Ether
Concen: 0.028 ug/l
RT: 2.123 min Scan# 11
Delta R.T. -0.012 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 74 Resp: 34
Ion Ratio Lower Upper
74 100
45 244.1 48.4 145.0#



#9
1,1,2-Trichlorotrifluoroethane
Concen: 1.894 ug/l
RT: 2.319 min Scan# 202
Delta R.T. 0.006 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09


Tgt Ion: 101 Resp: 2741
Ion Ratio Lower Upper
101 100
85 45.1 38.5 57.7
151 67.3 55.8 83.6








#10
Methyl Iodide
Concen: 0.011 ug/l
RT: 2.160 min Scan# 11
Delta R.T. -0.281 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:142 Resp: 21
Ion Ratio Lower Upper
142 100
127 0.0 41.8 62.8#
141 0.0 11.5 17.3#



#11
Tert butyl alcohol
Concen: 20.605 ug/l
RT: 2.947 min Scan# 305
Delta R.T. -0.049 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Tgt Ion: 59 Resp: 6039
Ion Ratio Lower Upper
59 100
57 0.3 8.8 13.2#





#12

1,1-Dichloroethene

Concen: 0.426 ug/l

RT: 2.306 min Scan# 200

Delta R.T. -0.000 min

Lab File: VX044252.D





Acq: 12 Dec 2024 15:09

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion:	96	Resp:	566
Ion	Ratio	Lower	Upper
96	100		
61	209.0	138.0	207.0#
98	67.3	50.0	75.0



#13

Acrolein

Concen: 0.099 ug/l




RT: 2.203 min Scan# 183


Delta R.T. -0.037 min

Lab File: VX044252.D

Acq: 12 Dec 2024 15:09

Tgt Ion:	56	Resp:	42
Ion	Ratio	Lower	Upper
56	100		
55	0.0	56.6	84.8#









#14
 Allyl chloride
 Concen: 0.041 ug/l
 RT: 2.666 min Scan# 21
 Delta R.T. 0.012 min
 Lab File: VX044252.D
 Acq: 12 Dec 2024 15:09




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 41 Resp: 90
 Ion Ratio Lower Upper
 41 100
 39 400.0 61.9 92.9#
 76 0.0 27.4 41.2#



#15
 Acrylonitrile
 Concen: 0.055 ug/l
 RT: 2.953 min Scan# 306
 Delta R.T. -0.116 min
 Lab File: VX044252.D
 Acq: 12 Dec 2024 15:09


Tgt Ion: 53 Resp: 45
 Ion Ratio Lower Upper
 53 100
 52 0.0 67.4 101.0#
 51 0.0 30.3 45.5#








#16
Acetone
Concen: 1.092 ug/l
RT: 2.380 min Scan# 214
Delta R.T. -0.012 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 43 Resp: 926
Ion Ratio Lower Upper
43 100
58 15.9 21.6 32.4#



#17
Carbon Disulfide
Concen: 0.076 ug/l
RT: 2.501 min Scan# 232
Delta R.T. -0.000 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Tgt Ion: 76 Resp: 192
Ion Ratio Lower Upper
76 100
78 0.0 7.4 11.0#





#18

Methyl Acetate

Concen: 0.199 ug/l

RT: 2.703 min Scan# 2

Delta R.T. -0.006 min


Lab File: VX044252.D

Acq: 12 Dec 2024 15:09

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion: 43 Resp: 436

Ion Ratio Lower Upper

43 100

74 0.0 17.7 26.5#



#19

Methyl tert-butyl Ether

Concen: 0.011 ug/l

RT: 3.123 min Scan# 334

Delta R.T. 0.006 min

Lab File: VX044252.D




Acq: 12 Dec 2024 15:09


Tgt Ion: 73 Resp: 55

Ion Ratio Lower Upper

73 100


57 0.0 16.2 24.4#








#20
Methylene Chloride
Concen: 0.303 ug/l
RT: 2.782 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion: 84 Resp: 482




Ion	Ratio	Lower	Upper
84	100		
49	82.9	104.2	156.4#
51	0.0	32.2	48.4#
86	56.4	50.9	76.3




#21
trans-1,2-Dichloroethene
Concen: 0.398 ug/l
RT: 3.081 min Scan# 327
Delta R.T. -0.000 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Tgt Ion: 96 Resp: 562

Ion	Ratio	Lower	Upper
96	100		
61	43.2	112.0	168.0#
98	32.5	48.8	73.2#









#23
 Vinyl Acetate
 Concen: 0.005 ug/l
 RT: 3.727 min Scan# 41
 Delta R.T. 0.006 min
 Lab File: VX044252.D
 Acq: 12 Dec 2024 15:09




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 43 Resp: 20
 Ion Ratio Lower Upper
 43 100
 86 0.0 8.2 12.2#



#24
 1,1-Dichloroethane
 Concen: 0.058 ug/l
 RT: 3.617 min Scan# 415
 Delta R.T. 0.012 min
 Lab File: VX044252.D
 Acq: 12 Dec 2024 15:09

Tgt Ion: 63 Resp: 161
 Ion Ratio Lower Upper
 63 100
 98 0.0 3.6 10.8#
 100 0.0 2.1 6.3#









#25
2-Butanone
Concen: 0.021 ug/l
RT: 4.428 min Scan# 54
Delta R.T. -0.134 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 25
Ion Ratio Lower Upper
43 100
72 0.0 20.2 30.4#



#27
cis-1,2-Dichloroethene
Concen: 1.505 ug/l
RT: 4.483 min Scan# 557
Delta R.T. -0.000 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Tgt Ion: 96 Resp: 2700
Ion Ratio Lower Upper
96 100
61 136.4 0.0 299.0
98 29.9 0.0 129.0





#29

Tetrahydrofuran

Concen: 0.025 ug/l

RT: 5.013 min Scan# 64

Delta R.T. 0.006 min

Lab File: VX044252.D

Acq: 12 Dec 2024 15:09

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 42 Resp: 19

Ion Ratio Lower Upper

42 100

72 0.0 37.0 55.6#

71 0.0 33.6 50.4#



#30

Chloroform

Concen: 0.173 ug/l

RT: 5.092 min Scan# 657

Delta R.T. 0.006 min

Lab File: VX044252.D




Acq: 12 Dec 2024 15:09


Tgt Ion: 83 Resp: 539

Ion Ratio Lower Upper

83 100


85 33.8 51.6 77.4#








#31
Cyclohexane
Concen: 1.571 ug/l
RT: 5.550 min Scan# 718
Delta R.T. 0.085 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 56 Resp: 3527
Ion Ratio Lower Upper
56 100
69 137.6 24.7 37.1#
84 84.3 71.0 106.6



#32
1,1,1-Trichloroethane
Concen: 0.007 ug/l
RT: 5.562 min Scan# 734
Delta R.T. 0.183 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Tgt Ion: 97 Resp: 19
Ion Ratio Lower Upper
97 100
99 430405.3 50.8 76.2#
61 22831.6 35.2 52.8#









#33
1,2-Dichloroethane-d4
Concen: 50.401 ug/l
RT: 5.952 min Scan# 798
Delta R.T. -0.000 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 65 Resp: 104813
Ion Ratio Lower Upper
65 100
67 50.2 0.0 101.6



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 930
Delta R.T. -0.000 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Tgt Ion: 114 Resp: 225924
Ion Ratio Lower Upper
114 100
63 21.9 0.0 40.6
88 16.3 0.0 31.8









#35
Dibromofluoromethane
Concen: 48.202 ug/l
RT: 5.379 min Scan# 705
Delta R.T. -0.006 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:113 Resp: 75423
Ion Ratio Lower Upper
113 100
111 99.8 80.4 120.6
192 16.5 13.3 19.9



#36
1,1-Dichloropropene
Concen: 5.613 ug/l
RT: 5.538 min Scan# 730
Delta R.T. -0.153 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09


Tgt Ion: 75 Resp: 11353
Ion Ratio Lower Upper
75 100
110 5.0 17.3 51.7#
77 0.0 25.4 38.0#








#37
Ethyl Acetate
Concen: 0.020 ug/l
RT: 4.782 min Scan# 606
Delta R.T. 0.061 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:	43	Resp:	50
Ion	Ratio	Lower	Upper
43	100		
61	0.0	9.8	14.8#
70	0.0	7.7	11.5#



#38
Carbon Tetrachloride
Concen: 6.369 ug/l
RT: 5.544 min Scan# 731
Delta R.T. -0.128 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Tgt Ion:	117	Resp:	13868
Ion	Ratio	Lower	Upper
117	100		
119	6.2	74.2	111.4#
121	0.0	23.0	34.6#





#39

Methylcyclohexane

Concen: 2.340 ug/l

RT: 7.117 min Scan# 91

Delta R.T. -0.256 min

Lab File: VX044252.D

Acq: 12 Dec 2024 15:09

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 83 Resp: 5657

Ion Ratio Lower Upper

83 100

55 0.0 60.7 91.1#

98 171.2 38.9 58.3#



#40

Benzene

Concen: 0.016 ug/l

RT: 6.044 min Scan# 813

Delta R.T. 0.012 min

Lab File: VX044252.D




Acq: 12 Dec 2024 15:09


Tgt Ion: 78 Resp: 98

Ion Ratio Lower Upper

78 100


77 0.0 19.3 28.9#








#41
Methacrylonitrile
Concen: 0.056 ug/l
RT: 4.910 min Scan# 61
Delta R.T. -0.012 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:	41	Resp:	75
Ion	Ratio	Lower	Upper
41	100		
39	237.3	44.2	66.4#
67	0.0	57.2	85.8#
52	0.0	24.6	37.0#



#42
1,2-Dichloroethane
Concen: 0.125 ug/l
RT: 6.098 min Scan# 822
Delta R.T. 0.012 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Tgt Ion:	62	Resp:	315
Ion	Ratio	Lower	Upper
62	100		
98	0.0	0.0	15.8









#43
Isopropyl Acetate
Concen: 0.005 ug/l
RT: 6.361 min Scan# 80
Delta R.T. 0.018 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 19
Ion Ratio Lower Upper
43 100
61 0.0 14.2 21.2#
87 0.0 10.2 15.2#



#44
Trichloroethene
Concen: 258.228 ug/l
RT: 7.123 min Scan# 990
Delta R.T. -0.000 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Tgt Ion:130 Resp: 395453
Ion Ratio Lower Upper
130 100
95 104.3 0.0 201.4









#47
Bromodichloromethane
Concen: 0.022 ug/l
RT: 7.836 min Scan# 1107
Delta R.T. 0.012 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 47
Ion Ratio Lower Upper
83 100
85 0.0 50.6 75.8#
127 0.0 7.3 10.9#



#48
Methyl methacrylate
Concen: 0.016 ug/l
RT: 7.665 min Scan# 1079
Delta R.T. -0.031 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Tgt Ion: 41 Resp: 31
Ion Ratio Lower Upper
41 100
69 0.0 62.4 93.6#
39 335.5 53.3 79.9#









#49
 1,4-Dioxane
 Concen: 1.781 ug/l
 RT: 7.683 min Scan# 1082
 Delta R.T. 0.012 min
 Lab File: VX044252.D
 Acq: 12 Dec 2024 15:09




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 88 Resp: 56
 Ion Ratio Lower Upper
 88 100
 43 0.0 27.1 40.7#
 58 0.0 41.0 61.4#



#50
 Toluene-d8
 Concen: 52.105 ug/l
 RT: 8.647 min Scan# 1240
 Delta R.T. -0.000 min
 Lab File: VX044252.D
 Acq: 12 Dec 2024 15:09

Tgt Ion: 98 Resp: 275060
 Ion Ratio Lower Upper
 98 100
 100 64.5 51.5 77.3









#51
4-Methyl-2-Pentanone
Concen: 0.016 ug/l
RT: 8.549 min Scan# 11
Delta R.T. -0.025 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 39
Ion Ratio Lower Upper
43 100
58 0.0 29.4 44.0#



#52
Toluene
Concen: 0.071 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Tgt Ion: 92 Resp: 272
Ion Ratio Lower Upper
92 100
91 146.3 137.0 205.6









#53
t-1,3-Dichloropropene
Concen: 1.006 ug/l
RT: 9.000 min Scan# 11
Delta R.T. 0.024 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 22
Ion Ratio Lower Upper
75 100
77 0.0 27.2 40.8#



#55
1,1,2-Trichloroethane
Concen: 0.164 ug/l
RT: 9.165 min Scan# 1325
Delta R.T. 0.012 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Tgt Ion: 97 Resp: 251
Ion Ratio Lower Upper
97 100
83 110.3 65.3 97.9#
85 88.7 44.6 67.0#
99 0.0 50.2 75.2#









#60
Dibromochloromethane
Concen: 2.674 ug/l
RT: 9.281 min Scan# 11
Delta R.T. -0.238 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:129 Resp: 2297
Ion Ratio Lower Upper
129 100
127 0.0 38.5 115.3#



#62
4-Bromofluorobenzene
Concen: 49.372 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Tgt Ion: 95 Resp: 91071
Ion Ratio Lower Upper
95 100
174 71.5 0.0 145.8
176 68.9 0.0 137.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1470
Delta R.T. 0.006 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:117 Resp: 196920
Ion Ratio Lower Upper
117 100
82 57.3 45.8 68.8
119 32.7 25.0 37.4



#64
Tetrachloroethene
Concen: 1.572 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.006 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Tgt Ion:164 Resp: 2054
Ion Ratio Lower Upper
164 100
166 99.4 100.4 150.6#
129 78.3 77.2 115.8
131 81.5 76.2 114.2









#65
Chlorobenzene
Concen: 0.020 ug/l
RT: 10.085 min Scan# 1475
Delta R.T. 0.006 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:112 Resp: 86
Ion Ratio Lower Upper
112 100
114 0.0 25.0 37.6#



#67
Ethyl Benzene
Concen: 0.046 ug/l
RT: 10.201 min Scan# 1495
Delta R.T. 0.012 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Tgt Ion: 91 Resp: 335
Ion Ratio Lower Upper
91 100
106 0.0 23.5 35.3#









#68
m/p-Xylenes
Concen: 0.161 ug/l
RT: 10.305 min Scan# 1511
Delta R.T. 0.006 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:106 Resp: 431
Ion Ratio Lower Upper
106 100
91 191.9 170.2 255.2



#69
o-Xylene
Concen: 0.037 ug/l
RT: 10.646 min Scan# 1568
Delta R.T. 0.006 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Tgt Ion:106 Resp: 100
Ion Ratio Lower Upper
106 100
91 346.0 112.3 337.0#









#70
Styrene
Concen: 0.005 ug/l
RT: 10.689 min Scan# 11
Delta R.T. 0.036 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:104 Resp: 23
Ion Ratio Lower Upper
104 100
78 0.0 44.6 67.0#
103 0.0 45.4 68.0#



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.006 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Tgt Ion:152 Resp: 82721
Ion Ratio Lower Upper
152 100
115 63.2 44.5 133.3
150 157.9 0.0 353.8





#73

Isopropylbenzene

Concen: 0.007 ug/l

RT: 10.976 min Scan# 1622

Delta R.T. 0.018 min

Lab File: VX044252.D

Acq: 12 Dec 2024 15:09

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:105 Resp: 46

Ion Ratio Lower Upper

105 100

120 0.0 13.1 39.3#



#74

N-ethyl acetate

Concen: 0.014 ug/l

RT: 10.963 min Scan# 1620

Delta R.T. 0.122 min

Lab File: VX044252.D

Acq: 12 Dec 2024 15:09

Tgt Ion: 43 Resp: 40




Ion Ratio Lower Upper


43 100

70 0.0 31.9 47.9#

55 0.0 17.9 26.9#

61 0.0 17.9 26.9#









#75
1,1,2-Tetrachloroethane
Concen: 0.073 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.134 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 160
Ion Ratio Lower Upper
83 100
131 74.4 5.1 15.4#
85 0.0 32.5 97.4#



#76
1,2,3-Trichloropropane
Concen: 27.914 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.159 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Tgt Ion: 75 Resp: 51556
Ion Ratio Lower Upper
75 100
77 0.9 20.9 62.7#









#77
Bromobenzene
Concen: 0.012 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:156 Resp: 19
Ion Ratio Lower Upper
156 100
77 0.0 82.6 247.8#
158 0.0 48.8 146.4#



#78
n-propylbenzene
Concen: 0.028 ug/l
RT: 11.311 min Scan# 1677
Delta R.T. 0.006 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Tgt Ion: 91 Resp: 198
Ion Ratio Lower Upper
91 100
120 0.0 11.0 33.0#





#79

2-Chlorotoluene

Concen: 0.044 ug/l

RT: 11.311 min Scan# 1

Delta R.T. -0.049 min

Lab File: VX044252.D

Acq: 12 Dec 2024 15:09

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 91 Resp: 198

Ion Ratio Lower Upper

91 100

126 0.0 16.4 49.1#



#80

1,3,5-Trimethylbenzene

Concen: 0.017 ug/l

RT: 11.457 min Scan# 1701

Delta R.T. 0.006 min

Lab File: VX044252.D




Acq: 12 Dec 2024 15:09


Tgt Ion:105 Resp: 91

Ion Ratio Lower Upper

105 100

120 50.5 23.8 71.4









#81
trans-1,4-Dichloro-2-butene
Concen: 76.642 ug/l
RT: 11.079 min Scan# 1
Delta R.T. 0.061 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 51556
Ion Ratio Lower Upper
75 100
53 0.0 93.6 140.4#
89 0.0 37.3 55.9#



#82
4-Chlorotoluene
Concen: 0.008 ug/l
RT: 11.463 min Scan# 1702
Delta R.T. 0.012 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Tgt Ion: 91 Resp: 42
Ion Ratio Lower Upper
91 100
126 97.6 14.5 43.5#





#83

tert-Butylbenzene

Concen: 0.022 ug/l

RT: 11.725 min Scan# 11

Delta R.T. 0.012 min

Lab File: VX044252.D





Acq: 12 Dec 2024 15:09

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion:119	Resp:	116
Ion	Ratio	Lower Upper
119	100	
91	16.4	30.9 92.5#
134	19.0	11.5 34.4



#84

1,2,4-Trimethylbenzene

Concen: 0.015 ug/l




RT: 11.762 min Scan# 1751


Delta R.T. 0.012 min

Lab File: VX044252.D

Acq: 12 Dec 2024 15:09

Tgt Ion:105	Resp:	78
Ion	Ratio	Lower Upper
105	100	
120	34.6	21.7 65.1





#85

sec-Butylbenzene

Concen: 0.012 ug/l

RT: 11.762 min Scan# 1751

Delta R.T. -0.128 min

Lab File: VX044252.D

Acq: 12 Dec 2024 15:09

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:105 Resp: 78

Ion Ratio Lower Upper

105 100

134 0.0 9.8 29.4#



#86

p-Isopropyltoluene

Concen: 0.017 ug/l

RT: 12.012 min Scan# 1792

Delta R.T. 0.006 min

Lab File: VX044252.D

Acq: 12 Dec 2024 15:09




Tgt Ion:119 Resp: 87


Ion Ratio Lower Upper

119 100

134 104.6 12.8 38.4#

91 260.9 13.0 39.0#





#87

1,3-Dichlorobenzene

Concen: 0.007 ug/l

RT: 11.981 min Scan# 1787

Delta R.T. 0.012 min

Lab File: VX044252.D

Acq: 12 Dec 2024 15:09

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:146 Resp: 20

Ion Ratio Lower Upper

146 100

111 190.0 21.4 64.2#

148 0.0 31.7 95.1#



#88

1,4-Dichlorobenzene

Concen: 0.113 ug/l

RT: 12.030 min Scan# 1795

Delta R.T. -0.012 min

Lab File: VX044252.D

Acq: 12 Dec 2024 15:09




Tgt Ion:146 Resp: 312


Ion Ratio Lower Upper

146 100

111 397.4 21.0 63.0#

148 138.8 32.3 96.8#





#89

n-Butylbenzene

Concen: 0.005 ug/l

RT: 12.372 min Scan# 11

Delta R.T. 0.043 min

Lab File: VX044252.D

Acq: 12 Dec 2024 15:09

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 91 Resp: 20

Ion Ratio Lower Upper

91 100

92 0.0 26.6 79.7#

134 0.0 12.6 37.8#



#90

Hexachloroethane

Concen: 0.052 ug/l

RT: 12.305 min Scan# 1840

Delta R.T. -0.232 min

Lab File: VX044252.D




Acq: 12 Dec 2024 15:09


Tgt Ion: 117 Resp: 40

Ion Ratio Lower Upper

117 100

201 0.0 31.4 94.2#









#95
Naphthalene
Concen: 0.069 ug/l
RT: 13.786 min Scan# 2083
Delta R.T. 0.012 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion:128 Resp: 413
Ion Ratio Lower Upper
128 100
127 0.0 10.5 15.7#
129 0.0 9.0 13.4#



#96
1,2,3-Trichlorobenzene
Concen: 0.013 ug/l
RT: 13.975 min Scan# 2114
Delta R.T. 0.012 min
Lab File: VX044252.D
Acq: 12 Dec 2024 15:09

Tgt Ion:180 Resp: 21
Ion Ratio Lower Upper
180 100
182 0.0 48.0 144.0#
145 0.0 17.7 53.1#

